

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.032 °C	29.8 °C	-0.232 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.631 %HR	35 %HR	4.369 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.854 %HR	53 %HR	3.146 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.460 %HR	73 %HR	1.540 %HR	± 5.0 %HR	± 1.9E+00 %HR