

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.032 °C         | 29.5 °C   | -0.532 °C     | ± 1.0 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.631 %HR        | 34 %HR    | 3.369 %HR     | ± 5.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 49.854 %HR        | 53 %HR    | 3.146 %HR     | ± 5.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 71.460 %HR        | 74 %HR    | 2.540 %HR     | ± 5.0 %HR            | ± 1.9E+00 %HR               |