

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|--------------------------|------------------------------------|------------|
| 1.005 mm                 | 1.010 mm         | 5 μm                 | ± 50.25 μm               | ±                                  | 5.8E-01 μm |
| 1.5 mm                   | 1.506 mm         | 6 μm                 | ± 75. μm                 | ±                                  | 5.8E-01 μm |
| 2 mm                     | 2.004 mm         | 4 μm                 | ± 100 μm                 | ±                                  | 5.8E-01 μm |
| 4 mm                     | 4.003 mm         | 3 μm                 | ± 200 μm                 | ±                                  | 5.8E-01 μm |
| 5 mm                     | 5.003 mm         | 3 μm                 | ± 250 μm                 | ±                                  | 5.8E-01 μm |