

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)	
52.65 μm	52.40 μm	-0.25 μm	± 1.6 μm	±	5.8E-02 μm
155.23 μm	155.0 μm	-0.23 μm	± 4.7 μm	±	6.2E-02 μm
257.12 μm	257.8 μm	0.68 μm	± 7.7 μm	±	7.0E-02 μm
511.72 μm	511.2 μm	-0.52 μm	± 15.4 μm	±	9.7E-02 μm