

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>
52.65 $\mu\text{m}$	52.98 $\mu\text{m}$	0.33 $\mu\text{m}$	$\pm$ 1.6 $\mu\text{m}$	$\pm$ 5.8E-02 $\mu\text{m}$
155.23 $\mu\text{m}$	155.0 $\mu\text{m}$	-0.23 $\mu\text{m}$	$\pm$ 4.7 $\mu\text{m}$	$\pm$ 6.2E-02 $\mu\text{m}$
257.12 $\mu\text{m}$	257.2 $\mu\text{m}$	0.08 $\mu\text{m}$	$\pm$ 7.7 $\mu\text{m}$	$\pm$ 7.0E-02 $\mu\text{m}$
511.72 $\mu\text{m}$	512.0 $\mu\text{m}$	0.28 $\mu\text{m}$	$\pm$ 15.4 $\mu\text{m}$	$\pm$ 9.7E-02 $\mu\text{m}$