

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.75	0.74925	1.020	± 1.8	± 2.9E+00 mm					
0 a 0.8	0.79915	1.120	± 1.8	± 2.9E+00 mm					
0 a 0.9	0.89905	1.230	± 1.9	± 2.9E+00 mm					
0 a 1	0.99895	1.330	± 2	± 2.9E+00 mm					
0 a 1.1	1.09885	1.430	± 2	± 2.9E+00 mm					
0 a 1.2	1.19875	1.540	± 2.1	± 2.9E+00 mm					
0 a 1.3	1.29865	1.640	± 2.2	± 2.9E+00 mm					
0 a 1.4	1.39855	1.750	± 2.3	± 2.9E+00 mm					
0 a 1.5	1.49845	1.850	± 2.4	± 2.9E+00 mm					
0 a 1.6	1.59835	1.960	± 2.4	± 2.9E+00 mm					
0 a 1.7	1.69825	2.060	± 2.5	± 2.9E+00 mm					
0 a 1.8	1.79815	2.170	± 2.6	± 2.9E+00 mm					
0 a 1.95	1.94805	2.280	± 2.7	± 2.9E+00 mm					
