

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
100.0905 µl	100 µl	-0.091 µl	± 0.8 µl	±	5.3E-02 µl
500.0516 µl	500 µl	-0.052 µl	± 4 µl	±	6.0E-02 µl
1000.003 µl	1000 µl	-0.003 µl	± 8 µl	±	8.2E-02 µl

