

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
9.9649 µl	10 µl	0.035 µl	± 0.12 µl	±	5.1E-02 µl
49.9751 µl	50 µl	0.025 µl	± 0.50 µl	±	5.2E-02 µl
100.121 µl	100 µl	-0.121 µl	± 0.80 µl	±	5.3E-02 µl