

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.423 °C	30.5 °C	0.077 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.609 %HR	31.4 %HR	1.791 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.640 %HR	50.9 %HR	1.260 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.090 %HR	73.9 %HR	-1.190 %HR	± 3.0 %HR	± 1.9E+00 %HR