

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) | |
|--------------------------|-------------------|----------------------|--------------------------|------------------------------------|-----------------------|
| 23.3 μm | 24 μm | 0.70 μm | $\pm 10 \mu\text{m}$ | \pm | 5.8E-01 μm |
| 51.6 μm | 50 μm | -1.60 μm | $\pm 10 \mu\text{m}$ | \pm | 5.8E-01 μm |
| 124.6 μm | 124 μm | -0.60 μm | $\pm 10 \mu\text{m}$ | \pm | 5.8E-01 μm |
| 250.3 μm | 249 μm | -1.30 μm | $\pm 18 \mu\text{m}$ | \pm | 5.8E-01 μm |
| 486.0 μm | 485 μm | -1.00 μm | $\pm 25 \mu\text{m}$ | \pm | 5.8E-01 μm |
| 963.0 μm | 962 μm | -1.00 μm | $\pm 39 \mu\text{m}$ | \pm | 6.0E-01 μm |