

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
57.0 μm	56.0 μm	-1.0 μm	$\pm 2.8 \mu\text{m}$	\pm	5.8E-01 μm
177.0 μm	176.0 μm	-1.0 μm	$\pm 8.8 \mu\text{m}$	\pm	5.8E-01 μm