

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49922	0.780	± 0.8	± 4.3E-01 mm					
0 a 1	0.99924	0.760	± 1	± 4.3E-01 mm					
0 a 2	1.99929	0.710	± 1.4	± 4.8E-01 mm					
0 a 4	3.99937	0.630	± 2.2	± 4.8E-01 mm					
0 a 6	5.99945	0.550	± 3	± 3.6E-01 mm					
0 a 10	9.99962	0.380	± 4.6	± 3.6E-01 mm					
0 a 15	14.99983	0.170	± 6.6	± 3.6E-01 mm					
0 a 18	17.99996	0.040	± 7.8	± 3.6E-01 mm					
0 a 21	21.00007	-0.070	± 9	± 5.5E-01 mm					
0 a 24	24.00018	-0.180	± 10.2	± 5.5E-01 mm					
0 a 30	30.00040	-0.400	± 12.6	± 5.5E-01 mm					
0 a 35	35.00058	-0.580	± 14.6	± 5.5E-01 mm					
0 a 40	40.00074	-0.740	± 16.6	± 5.5E-01 mm					
0 a 45	45.00090	-0.900	± 18.6	± 6.3E-01 mm					
0 a 50	50.00104	-1.040	± 20.6	± 6.3E-01 mm					
