

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
0 °C	-0.201 °C	0.1 °C	0.301 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.787 °C	9.9 °C	0.113 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.804 °C	19.7 °C	-0.104 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.932 °C	29.6 °C	-0.332 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.930 °C	39.5 °C	-0.430 °C	± 0.5 °C	± 4.0E-01 °C