

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.201 °C	0.2 °C	0.401 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.787 °C	10.2 °C	0.413 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.804 °C	19.9 °C	0.096 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.932 °C	29.8 °C	-0.132 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.930 °C	39.7 °C	-0.230 °C	± 0.5 °C	± 4.0E-01 °C