

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.032 °C	30.3 °C	0.268 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.325 %HR	33 %HR	-2.325 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.231 %HR	52 %HR	2.769 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.558 %HR	75 %HR	5.442 %HR	± 5.0 %HR	± 1.9E+00 %HR