

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.032 °C	29.4 °C	-0.632 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.325 %HR	37 %HR	1.675 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.231 %HR	52 %HR	2.769 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.558 %HR	74 %HR	4.442 %HR	± 5.0 %HR	± 1.9E+00 %HR