

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C | 30.032 °C | 30.5 °C | 0.468 °C | ± 1.0 °C | ± 4.0E-01 °C |
| 35 %HR | 35.325 %HR | 46 %HR | 10.675 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 49.231 %HR | 62 %HR | 12.769 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 70 %HR | 69.558 %HR | 83 %HR | 13.442 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |