

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09970	0.300	± 1.2	± 4.3E-01 mm					
0 a 0.15	0.14961	0.390	± 1.3	± 4.3E-01 mm					
0 a 0.2	0.19956	0.440	± 1.3	± 4.3E-01 mm					
0 a 0.25	0.24952	0.480	± 1.4	± 4.3E-01 mm					
0 a 0.3	0.29946	0.540	± 1.4	± 4.3E-01 mm					
0 a 0.35	0.34947	0.530	± 1.4	± 4.3E-01 mm					
0 a 0.4	0.39939	0.610	± 1.5	± 4.3E-01 mm					
0 a 0.45	0.44937	0.630	± 1.5	± 4.3E-01 mm					
0 a 0.5	0.49929	0.710	± 1.6	± 4.3E-01 mm					
0 a 0.55	0.54920	0.800	± 1.6	± 4.3E-01 mm					
0 a 0.6	0.59919	0.810	± 1.6	± 4.3E-01 mm					
