

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09982	0.180	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.14974	0.260	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.19973	0.270	± 0.6	± 4.3E-01 mm					
0 a 0.25	0.24965	0.350	± 0.7	± 4.3E-01 mm					
0 a 0.3	0.29961	0.390	± 0.7	± 4.3E-01 mm					
0 a 0.35	0.34960	0.400	± 0.7	± 4.3E-01 mm					
0 a 0.4	0.39951	0.490	± 0.7	± 4.3E-01 mm					
0 a 0.45	0.44947	0.530	± 0.7	± 4.3E-01 mm					
0 a 0.5	0.49937	0.630	± 0.8	± 4.3E-01 mm					
0 a 0.55	0.54932	0.680	± 0.8	± 4.3E-01 mm					
0 a 0.6	0.59926	0.740	± 0.8	± 4.3E-01 mm					
