

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09978	0.220	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.14973	0.270	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.19966	0.340	± 0.6	± 4.3E-01 mm					
0 a 0.25	0.24963	0.370	± 0.7	± 4.3E-01 mm					
0 a 0.3	0.29957	0.430	± 0.7	± 4.3E-01 mm					
0 a 0.35	0.34957	0.430	± 0.7	± 4.3E-01 mm					
0 a 0.4	0.39947	0.530	± 0.7	± 4.3E-01 mm					
0 a 0.45	0.44937	0.630	± 0.7	± 4.3E-01 mm					
0 a 0.5	0.49936	0.640	± 0.8	± 4.3E-01 mm					
0 a 0.55	0.54929	0.710	± 0.8	± 4.3E-01 mm					
0 a 0.6	0.59928	0.720	± 0.8	± 4.3E-01 mm					
