

Clase de Exactitud (Accuracy Class): Clase II (Class II)

---

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09986	0.140	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.14979	0.210	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.19976	0.240	± 0.6	± 4.3E-01 mm					
0 a 0.25	0.24969	0.310	± 0.7	± 4.3E-01 mm					
0 a 0.3	0.29964	0.360	± 0.7	± 4.3E-01 mm					
0 a 0.35	0.34960	0.400	± 0.7	± 4.3E-01 mm					
0 a 0.4	0.39956	0.440	± 0.7	± 4.3E-01 mm					
0 a 0.45	0.44946	0.540	± 0.7	± 4.3E-01 mm					
0 a 0.5	0.49942	0.580	± 0.8	± 4.3E-01 mm					
0 a 0.55	0.54936	0.640	± 0.8	± 4.3E-01 mm					
0 a 0.6	0.59927	0.730	± 0.8	± 4.3E-01 mm					

---