

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1.00 mm	1.01 mm	10 µm	± 10 µm	± 5.8E+00 µm
1.25 mm	1.27 mm	20 µm	± 10 µm	± 5.8E+00 µm
1.50 mm	1.52 mm	20 µm	± 10 µm	± 5.8E+00 µm
2.00 mm	2.02 mm	20 µm	± 10 µm	± 5.8E+00 µm
3.00 mm	3.02 mm	20 µm	± 10 µm	± 5.8E+00 µm
4.00 mm	4.02 mm	20 µm	± 10 µm	± 5.8E+00 µm
5.00 mm	5.02 mm	20 µm	± 10 µm	± 5.8E+00 µm
8.00 mm	8.02 mm	20 µm	± 10 µm	± 5.8E+00 µm
10.00 mm	10.02 mm	20 µm	± 10 µm	± 5.8E+00 µm