

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
2.5 RPM	2.50 RPM	0.0 RPM	± 0.10 RPM	±	5.8E-02 RPM
35 RPM	35.0 RPM	0.0 RPM	± 0.11 RPM	±	5.8E-02 RPM
150 RPM	150.0 RPM	0.0 RPM	± 0.13 RPM	±	5.8E-02 RPM
580 RPM	579.9 RPM	-0.1 RPM	± 0.22 RPM	±	5.8E-02 RPM
800 RPM	799.9 RPM	-0.1 RPM	± 0.26 RPM	±	5.8E-02 RPM
1100 RPM	1099.9 RPM	-0.1 RPM	± 0.32 RPM	±	5.8E-02 RPM
99999 RPM	99997 RPM	-2.0 RPM	± 21 RPM	±	9.3E-01 RPM