

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
2:15:00 PM	2.4	2.30	-0.1
2:15:30 PM	2.4	2.30	-0.1
2:16:00 PM	2.4	2.30	-0.1
2:16:30 PM	2.4	2.30	-0.1
2:17:00 PM	2.4	2.30	-0.1
2:17:30 PM	2.4	2.20	-0.2
2:18:00 PM	2.3	2.20	-0.1
2:18:30 PM	2.3	2.20	-0.1
2:19:00 PM	2.3	2.20	-0.1
2:19:30 PM	2.3	2.30	0.0
2:20:00 PM	2.3	2.30	0.0
2:20:30 PM	2.3	2.30	0.0
2:21:00 PM	2.3	2.30	0.0
2:21:30 PM	2.3	2.30	0.0
2:22:00 PM	2.4	2.30	-0.1
2:22:30 PM	2.4	2.30	-0.1
2:23:00 PM	2.4	2.40	0.0
2:23:30 PM	2.4	2.40	0.0
2:24:00 PM	2.4	2.40	0.0
2:24:30 PM	2.4	2.40	0.0
2:25:00 PM	2.3	2.40	0.1
2:25:30 PM	2.3	2.40	0.1
2:26:00 PM	2.3	2.3	0.0
2:26:30 PM	2.3	2.3	0.0
2:27:00 PM	2.3	2.3	0.0
2:27:30 PM	2.3	2.3	0.0
2:28:00 PM	2.2	2.3	0.1
2:28:30 PM	2.2	2.3	0.1
2:29:00 PM	2.2	2.2	0.0
2:29:30 PM	2.1	2.2	0.1
Promedio	2.3	2.3	
Desviación	0.077	0.064	
Max	2.4	2.4	
Min	2.1	2.2	

Tabla 2.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
2:35:00 PM	7.4	7.5	0.1
2:35:30 PM	7.4	7.5	0.1
2:36:00 PM	7.5	7.5	0.0
2:36:30 PM	7.5	7.6	0.1
2:37:00 PM	7.5	7.6	0.1
2:37:30 PM	7.5	7.6	0.1
2:38:00 PM	7.6	7.7	0.1
2:38:30 PM	7.6	7.7	0.1
2:39:00 PM	7.6	7.7	0.1
2:39:30 PM	7.6	7.7	0.1
2:40:00 PM	7.6	7.7	0.1
2:40:30 PM	7.6	7.8	0.2
2:41:00 PM	7.7	7.8	0.1
2:41:30 PM	7.7	7.8	0.1
2:42:00 PM	7.8	7.9	0.1
2:42:30 PM	7.8	7.9	0.1
2:43:00 PM	7.9	7.8	-0.1
2:43:30 PM	7.9	7.8	-0.1
2:44:00 PM	7.8	7.7	-0.1
2:44:30 PM	7.8	7.7	-0.1
2:45:00 PM	7.6	7.7	0.1
2:45:30 PM	7.5	7.6	0.1
2:46:00 PM	7.5	7.6	0.1
2:46:30 PM	7.5	7.6	0.1
2:47:00 PM	7.5	7.6	0.1
2:47:30 PM	7.6	7.7	0.1
2:48:00 PM	7.6	7.7	0.1
2:48:30 PM	7.6	7.7	0.1
2:49:00 PM	7.7	7.8	0.1
2:49:30 PM	7.6	7.8	0.2
Promedio	7.6	7.7	
Desviación	0.134	0.108	
Max	7.9	7.9	
Min	7.4	7.5	

