



CERTIFICADO DE CALIBRACIÓN / CALIBRATION CERTIFICATE




Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
10:45:00 AM	2.3	2.20	-0.1
10:45:30 AM	2.3	2.20	-0.1
10:46:00 AM	2.3	2.20	-0.1
10:46:30 AM	2.4	2.20	-0.2
10:47:00 AM	2.4	2.30	-0.1
10:47:30 AM	2.4	2.30	-0.1
10:48:00 AM	2.4	2.30	-0.1
10:48:30 AM	2.5	2.30	-0.2
10:49:00 AM	2.5	2.40	-0.1
10:49:30 AM	2.5	2.40	-0.1
10:50:00 AM	2.5	2.40	-0.1
10:50:30 AM	2.5	2.40	-0.1
10:51:00 AM	2.5	2.40	-0.1
10:51:30 AM	2.5	2.40	-0.1
10:52:00 AM	2.3	2.40	0.1
10:52:30 AM	2.3	2.50	0.2
10:53:00 AM	2.3	2.50	0.2
10:53:30 AM	2.4	2.50	0.1
10:54:00 AM	2.4	2.50	0.1
10:54:30 AM	2.4	2.50	0.1
10:55:00 AM	2.3	2.40	0.1
10:55:30 AM	2.3	2.40	0.1
10:56:00 AM	2.3	2.40	0.1
10:56:30 AM	2.2	2.3	0.1
10:57:00 AM	2.2	2.3	0.1
10:57:30 AM	2.2	2.3	0.1
10:58:00 AM	2.2	2.3	0.1
10:58:30 AM	2.2	2.3	0.1
10:59:00 AM	2.1	2.2	0.1
10:59:30 AM	2.1	2.2	0.1
Promedio	2.3	2.3	
Desviación	0.122	0.101	
Max	2.5	2.5	
Min	2.1	2.2	

Tabla 2.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
11:15:00 AM	10.2	10.1	-0.1
11:15:30 AM	10.0	9.8	-0.2
11:16:00 AM	9.9	9.8	-0.1
11:16:30 AM	9.7	9.6	-0.1
11:17:00 AM	9.5	9.3	-0.2
11:17:30 AM	9.1	8.8	-0.3
11:18:00 AM	8.8	8.5	-0.3
11:18:30 AM	8.5	8.2	-0.3
11:19:00 AM	8.2	8.0	-0.2
11:19:30 AM	8.0	7.9	-0.1
11:20:00 AM	7.8	7.8	0.0
11:20:30 AM	7.7	7.8	0.1
11:21:00 AM	7.7	7.8	0.1
11:21:30 AM	7.7	7.7	0.0
11:22:00 AM	7.6	7.7	0.1
11:22:30 AM	7.6	7.7	0.1
11:23:00 AM	7.7	7.7	0.0
11:23:30 AM	7.7	7.6	-0.1
11:24:00 AM	7.8	7.6	-0.2
11:24:30 AM	7.8	7.6	-0.2
11:25:00 AM	7.8	7.6	-0.2
11:25:30 AM	7.9	7.7	-0.2
11:26:00 AM	7.9	7.7	-0.2
11:26:30 AM	8.0	7.7	-0.3
11:27:00 AM	8.0	7.8	-0.2
11:27:30 AM	8.0	7.8	-0.2
11:28:00 AM	8.0	7.8	-0.2
11:28:30 AM	7.5	7.6	0.1
11:29:00 AM	7.5	7.5	0.0
11:29:30 AM	7.4	7.6	0.2
Promedio	8.2	8.1	
Desviación	0.829	0.781	
Max	10.2	10.1	
Min	7.4	7.5	

Curva de Calibración - TEMPERATURA 2 °C



Curva de Calibración - TEMPERATURA 8 °C

