



CERTIFICADO DE CALIBRACIÓN / CALIBRATION CERTIFICATE



Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
9:35:00 AM	10.2	10.1	-0.1
9:35:30 AM	10.0	9.8	-0.2
9:36:00 AM	9.9	9.8	-0.1
9:36:30 AM	9.7	9.6	-0.1
9:37:00 AM	9.5	9.3	-0.2
9:37:30 AM	9.1	8.8	-0.3
9:38:00 AM	8.8	8.5	-0.3
9:38:30 AM	8.5	8.2	-0.3
9:39:00 AM	8.2	8.0	-0.2
9:49:30 AM	8.0	7.9	-0.1
9:40:00 AM	7.8	7.8	0.0
9:40:30 AM	7.7	7.8	0.1
9:41:00 AM	7.7	7.8	0.1
9:41:30 AM	7.7	7.7	0.0
9:42:00 AM	7.6	7.7	0.1
9:42:30 AM	7.6	7.7	0.1
9:43:00 AM	7.7	7.7	0.0
9:43:30 AM	7.7	7.6	-0.1
9:44:00 AM	7.8	7.6	-0.2
9:44:30 AM	7.8	7.6	-0.2
9:45:00 AM	7.8	7.6	-0.2
9:45:30 AM	7.9	7.7	-0.2
9:46:00 AM	7.9	7.7	-0.2
9:46:30 AM	8.0	7.7	-0.3
9:47:00 AM	8.0	7.8	-0.2
9:47:30 AM	8.0	7.8	-0.2
9:48:00 AM	8.0	7.8	-0.2
9:48:30 AM	7.5	7.6	0.1
9:49:00 AM	7.5	7.5	0.0
9:40:30 AM	7.4	7.6	0.2
Promedio	8.2	8.1	
Desviación	0.829	0.781	
Max	10.2	10.1	
Min	7.4	7.5	

