

Tabla 1.0: Lecturas de los sensores durante desplazamiento en la zona de trabajo a 100 °C, Bloque lado derecho

	Tiempo (hh:mm:ss)	Lectura del Patrón @ 100 °C				
		PC	P1	P2	P3	P4
1	10:54:56	100	91.7	90.9	92.1	91.3
2	10:55:56	100	91.7	90.9	92.2	91.6
3	10:56:56	100	91.7	91.0	92.2	91.6
4	10:57:56	100	91.7	90.9	92.3	91.6
5	10:58:56	100	91.8	91.1	92.3	91.6
6	10:59:56	100	91.7	91.0	92.2	91.6
7	11:00:56	100	91.7	90.9	92.2	91.6
8	11:01:56	100	91.8	91.0	92.3	91.6
9	11:02:56	100	91.7	91.0	92.2	91.6
10	11:03:56	100	91.7	90.9	92.2	91.6
11	11:04:56	100	91.8	91.0	92.2	91.6
12	11:05:56	100	91.7	91.0	92.2	91.6
13	11:06:56	100	91.8	91.0	92.2	91.6
14	11:07:56	100	91.7	90.9	92.2	91.5
15	11:08:56	100	91.7	91.0	92.2	91.6
16	11:09:56	100	91.7	90.9	92.2	91.5
17	11:10:56	100	91.7	90.9	92.2	91.5
18	11:11:56	100	91.7	90.9	92.1	91.5
19	11:12:56	100	91.7	90.9	92.3	91.5
20	11:13:56	100	91.6	90.9	92.1	91.3
21	11:14:56	100	91.6	90.8	92.1	91.2
22	11:15:56	100	91.7	90.9	92.2	91.3
23	11:16:56	100	91.6	90.8	92.1	91.2
24	11:17:56	100	91.7	90.9	92.2	91.3
Promedio		100	91.70	90.93	92.19	91.49
Desviación		0	0.05	0.06	0.06	0.15
Max		100	91.79	91.05	92.28	91.64
Min		100	91.57	90.81	92.05	91.19
ΔT Est.		0.45				
ΔT Radial		0.77				
ΔT Axial		0.49				

Fin de Certificado (End of Certificate)

Tabla 2.0: Lecturas de los sensores durante desplazamiento en la zona de trabajo 150 °C, Bloque lado izquierdo

	Tiempo (hh:mm:ss)	Lectura del Patrón °C @ 150 °C				
		PC	P1	P2	P3	P4
1	2:00:00 PM	150	146.2	146.1	146.8	148.1
2	2:01:00 PM	150	146.0	146.0	146.7	148.0
3	2:02:00 PM	150	146.0	146.0	146.8	148.0
4	2:03:00 PM	150	145.9	146.0	146.6	147.9
5	2:04:00 PM	150	145.8	145.8	146.4	147.8
6	2:05:00 PM	150	145.8	145.8	146.4	147.8
7	2:06:00 PM	150	145.7	145.8	146.4	147.8
8	2:07:00 PM	150	145.8	145.8	146.4	147.7
9	2:08:00 PM	150	145.7	145.7	146.3	147.7
10	2:09:00 PM	150	145.7	145.7	146.2	147.7
11	2:10:00 PM	150	145.6	145.6	146.2	147.6
12	2:11:00 PM	150	145.6	145.7	146.2	147.6
13	2:12:00 PM	150	145.6	145.7	146.2	147.6
14	2:13:00 PM	150	145.6	145.7	146.2	147.6
15	2:14:00 PM	150	145.6	145.6	146.3	147.7
16	2:15:00 PM	150	145.5	145.6	146.1	147.5
17	2:16:00 PM	150	145.5	145.4	146.1	147.4
18	2:17:00 PM	150	145.5	145.5	146.1	147.5
19	2:18:00 PM	150	145.5	145.6	146.2	147.5
20	2:19:00 PM	150	145.5	145.5	146.1	147.4
21	2:20:00 PM	150	145.4	145.5	146.1	147.5
22	2:21:00 PM	150	145.3	145.3	146.0	147.3
23	2:22:00 PM	150	145.2	145.3	145.9	147.4
24	2:23:00 PM	150	145.3	145.3	146.0	147.4
Promedio		150.00	145.64	145.66	146.29	147.65
Desviación		0.00	0.23	0.23	0.25	0.22
Max		150.00	146.16	146.13	146.84	148.11
Min		150.00	145.24	145.29	145.91	147.26
ΔT Est.		0.93				
ΔT Radial		0.02				
ΔT Axial		0.65				

Fin de Certificado (End of Certificate)