

| Certificado No.: 62228 |                   | Equipo (Instrument): Termómetro Digital |               |                        |        |                             |
|------------------------|-------------------|-----------------------------------------|---------------|------------------------|--------|-----------------------------|
| Fecha de Calibración:  |                   | 2025-10-15                              |               | Marca (Brand): ELITECH |        |                             |
| Nominal (Nominal)      | Patrón (Standard) | UBP (UUT)                               | Error (Error) | Exactitud (Accuracy)   |        | Incertidumbre (Uncertainty) |
| -30 °C                 | -30.321 °C        | -30.1 °C                                | 0.221 °C      | ±                      | 0.5 °C | ± 2.9E+00 °C                |
| -15 °C                 | -15.041 °C        | -15.3 °C                                | -0.259 °C     | ±                      | 0.5 °C | ± 2.9E+00 °C                |
| 5 °C                   | 5.079 °C          | 5.0 °C                                  | -0.079 °C     | ±                      | 0.5 °C | ± 2.9E+00 °C                |