

Certificado No.:

66977

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-10

TRACKLINK

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
2:03:00 PM	15.40	15.10	-0.30
2:04:00 PM	15.20	15.10	-0.10
2:05:00 PM	15.10	14.80	-0.30
2:06:00 PM	15.00	14.70	-0.30
2:07:00 PM	15.00	14.70	-0.30
2:08:00 PM	14.80	14.40	-0.40
2:09:00 PM	14.80	14.20	-0.60
2:10:00 PM	14.70	14.00	-0.70
2:11:00 PM	14.20	13.80	-0.40
2:12:00 PM	14.00	13.40	-0.60
2:13:00 PM	14.00	13.60	-0.40
2:14:00 PM	13.80	12.80	-1.00
2:15:00 PM	13.40	12.90	-0.50
2:16:00 PM	13.70	13.50	-0.20
2:17:00 PM	13.00	12.00	-1.00
2:18:00 PM	13.00	12.40	-0.60
2:19:00 PM	12.80	12.50	-0.30
2:20:00 PM	12.80	12.60	-0.20
2:21:00 PM	12.40	12.80	0.40
2:22:00 PM	12.10	12.90	0.80
2:23:00 PM	12.00	13.10	1.10
2:24:00 PM	12.00	13.20	1.20
2:25:00 PM	11.80	12.80	1.00
2:26:00 PM	11.90	12.50	0.60
2:27:00 PM	11.40	11.90	0.50
2:28:00 PM	11.50	11.50	0.00
2:29:00 PM	11.60	11.40	-0.20
2:30:00 PM	11.90	11.50	-0.40
2:31:00 PM	12.00	12.10	0.10
2:32:00 PM	12.40	12.50	0.10
Promedio	13.26	13.16	
Desviación	1.318	1.090	
Max	15.40	15.10	
Min	11.40	11.40	

Certificado No.:

66977

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-10

TRACKLINK

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
2:03:00 PM	70.50	72.50	2.00
2:04:00 PM	70.60	72.40	1.80
2:05:00 PM	70.20	72.10	1.90
2:06:00 PM	70.10	72.00	1.90
2:07:00 PM	70.00	72.00	2.00
2:08:00 PM	69.50	71.80	2.30
2:09:00 PM	69.80	71.90	2.10
2:10:00 PM	69.50	71.80	2.30
2:11:00 PM	69.40	71.50	2.10
2:12:00 PM	69.20	71.40	2.20
2:13:00 PM	69.00	71.10	2.10
2:14:00 PM	69.00	71.00	2.00
2:15:00 PM	68.70	71.00	2.30
2:16:00 PM	68.50	70.80	2.30
2:17:00 PM	68.20	70.90	2.70
2:18:00 PM	68.00	70.50	2.50
2:19:00 PM	67.80	70.40	2.60
2:20:00 PM	67.90	70.10	2.20
2:21:00 PM	67.50	70.00	2.50
2:22:00 PM	67.40	70.00	2.60
2:23:00 PM	67.10	69.80	2.70
2:24:00 PM	66.80	69.90	3.10
2:25:00 PM	66.90	69.70	2.80
2:26:00 PM	66.50	69.50	3.00
2:27:00 PM	66.40	69.40	3.00
2:28:00 PM	66.20	69.30	3.10
2:29:00 PM	66.10	69.00	2.90
2:30:00 PM	66.00	69.00	3.00
2:31:00 PM	65.80	68.90	3.10
2:32:00 PM	66.10	68.20	2.10
Promedio	68.16	70.60	
Desviación	1.528	1.186	
Max	70.60	72.50	
Min	65.80	68.90	

Certificado No.:

66977

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-10

TRACKLINK

