

CERTIFICADO DE CALIBRACIÓN / CALIBRATION CERTIFICATE

INNOVATECIS CIA LTDA

General José María Guerrero N69-170 y Alfonso del Hierro

Quito, Ecuador

(+593) 02 6040 607

innovatec@innovatec.com.ec

Certificado No. (Certificate #): 57465

Fecha de Recepción (Reception Date): 2025-05-29

Fecha de Calibración (Calibration Date): 2025-06-02

Próxima Fecha de Calibración (Calibration Due): 2026-06-02

Fecha de Emisión (Emission Date): 2025-06-05

Cliente (Client): BAXTER HEALTHCARE PANAMA, S.A

Dirección (Address): Corregimiento de San Francisco, Urbanización Punta Pacifica, Edificio Torre de las Américas, Oficina 2703 C.

Información del Instrumento (Instrument Information)

Equipo (Instrument):	Jeringa	Intervalo de Medición (Range):	100 µl	Material (Material):	Vidrio boro-silicato
Marca (Brand):	Hamilton				
Modelo (Model):	81020	División de escala: (Resolution)	1 µl	Ubicación (Location):	Laboratorio
Serie (Serial #):	INN-35953				
Código (Code):	SYP003	Lugar de Calibración (Place of Calibration):	Lab. INNOVATEC / INNOVATEC's Lab.		

Datos de Calibración (Calibration Info)
Procedimiento (Procedure): INN-PC-30

Condiciones Ambientales (Environmental Conditions)
Temperatura (Temp): (21.7 °C a 21.9 °C)

Humedad (Humidity): (70 %HR a 69 %HR)

Trazabilidad (Traceability Info)

Patrón (Standard)	Marca (Brand)	Cert. #	Última Calibración (Last Cal.)	Período (Period)
Balanza Analítica	Radwag	55581	2025-04-02	1 año

Resultados (Results)

UBP (UUT) µl	Patrón (Standard) µl	Error (Error) µl	Incertidumbre (Uncertainty)
10 As Found	10.5383	-0.5383	± 2.0 µl
50 As Found	49.6033	0.3967	± 2.5 µl
100 As Found	99.9485	0.0515	± 2.1 µl
10 As Left	10.5383	-0.5383	± 2.0 µl
50 As Left	49.6033	0.3967	± 2.5 µl
100 As Left	99.9485	0.0515	± 2.1 µl

El presente Certificado de Calibración posee la trazabilidad en esta magnitud hacia el Patrón Nacional, a través de la realización de la unidad de medida en el NPL, NIST, o otro Laboratorio Nacional reconocido al Sistema Internacional de Medidas. La calibración fue realizada bajo un Sistema de Gestión de Laboratorio conforme a la Norma ISO/IEC 17025:2017. Los resultados y su incertidumbre reportada con un nivel de confianza de K=2, 95% son relacionados a este instrumento y en el tiempo que se realizó las medidas. Este Laboratorio no se responsabiliza de los perjuicios que pueda ocasionar el uso inadecuado del instrumento calibrado. La reproducción parcial es prohibida, la reproducción total deberá hacerse con la autorización escrita aprobada por INNOVATEC Industrial Solutions. *This Certificate of Calibration provides traceability of measurement to the National Standard, through units of measurement realized at the NPL, NIST or other recognized National Standard Laboratories to the International System of Units. The calibration was performed under a Laboratory Management System in accordance with the ISO/IEC 17025:2017 Standard. The results and the reported uncertainty at a confidence level of K=2, 95% are related only to this instrument and at the time of measurement. This Laboratory is not responsible for any damages that may result from improper use of the calibrated instrument. Partial reproduction is forbidden, the total reproduction must have an approved written authorization by INNOVATEC Industrial Solutions.*

Comentarios: El equipo o instrumento bajo calibración cumple y fue encontrado dentro de tolerancia.

Comments

Calibrado por: Ing. Rubén Ortega

Calibrated by:

Aprobado por:

Approved by:


Fin de Certificado (End of Certificate)