



Certificado No.:

61073

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-08-28

QUECKLINC

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
9:05:00 AM	22.20	22.00	-0.20
9:06:00 AM	22.00	21.75	-0.25
9:07:00 AM	21.80	21.75	-0.05
9:08:00 AM	21.70	21.50	-0.20
9:09:00 AM	21.60	21.50	-0.10
9:10:00 AM	21.50	21.50	0.00
9:11:00 AM	21.50	21.50	0.00
9:12:00 AM	21.50	21.25	-0.25
9:13:00 AM	21.40	21.25	-0.15
9:14:00 AM	21.40	21.25	-0.15
9:15:00 AM	21.40	21.25	-0.15
9:16:00 AM	21.40	21.25	-0.15
9:17:00 AM	21.30	21.25	-0.05
9:18:00 AM	21.30	21.25	-0.05
9:19:00 AM	21.30	21.25	-0.05
9:20:00 AM	21.30	21.25	-0.05
9:21:00 AM	21.30	21.25	-0.05
9:22:00 AM	21.30	21.25	-0.05
9:23:00 AM	21.30	21.25	-0.05
9:24:00 AM	21.30	21.25	-0.05
9:25:00 AM	21.40	21.25	-0.15
9:26:00 AM	21.40	21.25	-0.15
9:27:00 AM	21.40	21.25	-0.15
9:28:00 AM	21.40	21.25	-0.15
9:29:00 AM	21.40	21.25	-0.15
9:30:00 AM	21.50	21.25	-0.25
9:31:00 AM	21.50	21.25	-0.25
9:32:00 AM	21.50	21.25	-0.25
9:33:00 AM	21.50	21.25	-0.25
9:34:00 AM	21.50	21.25	-0.25
<b>Promedio</b>	<b>21.48</b>	<b>21.34</b>	
<b>Desviación</b>	<b>0.210</b>	<b>0.191</b>	
<b>Max</b>	<b>22.20</b>	<b>22.00</b>	
<b>Min</b>	<b>21.30</b>	<b>21.25</b>	

Certificado No.:

61073

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-08-28

QUECKLINC

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
9:05:00 AM	40.50	40	-0.50
9:06:00 AM	40.50	40	-0.50
9:07:00 AM	40.60	40	-0.60
9:08:00 AM	40.60	40	-0.60
9:09:00 AM	40.60	39	-1.60
9:10:00 AM	40.70	39	-1.70
9:11:00 AM	40.80	39	-1.80
9:12:00 AM	40.80	39	-1.80
9:13:00 AM	40.80	39	-1.80
9:14:00 AM	40.80	39	-1.80
9:15:00 AM	40.90	39	-1.90
9:16:00 AM	40.90	39	-1.90
9:17:00 AM	41.00	39	-2.00
9:18:00 AM	41.00	39	-2.00
9:19:00 AM	41.00	39	-2.00
9:20:00 AM	41.00	39	-2.00
9:21:00 AM	41.00	39	-2.00
9:22:00 AM	41.10	39	-2.10
9:23:00 AM	41.10	39	-2.10
9:24:00 AM	41.10	39	-2.10
9:25:00 AM	41.10	39	-2.10
9:26:00 AM	41.20	39	-2.20
9:27:00 AM	41.10	39	-2.10
9:28:00 AM	41.10	39	-2.10
9:29:00 AM	41.20	39	-2.20
9:30:00 AM	41.10	39	-2.10
9:31:00 AM	41.10	39	-2.10
9:32:00 AM	41.10	39	-2.10
9:33:00 AM	41.10	39	-2.10
9:34:00 AM	41.00	39	-2.00
<b>Promedio</b>	<b>40.93</b>	<b>39.13</b>	
<b>Desviación</b>	<b>0.210</b>	<b>0.346</b>	
<b>Max</b>	<b>41.20</b>	<b>40.00</b>	
<b>Min</b>	<b>40.50</b>	<b>39.00</b>	

