

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
2:15:00 PM	20.30	20.10	-0.20
2:16:00 PM	20.10	20.20	0.10
2:17:00 PM	20.20	20.50	0.30
2:18:00 PM	20.40	20.60	0.20
2:19:00 PM	20.50	20.80	0.30
2:20:00 PM	20.10	20.90	0.80
2:21:00 PM	20.00	20.40	0.40
2:22:00 PM	19.90	20.10	0.20
2:23:00 PM	19.50	20.00	0.50
2:24:00 PM	19.50	20.00	0.50
2:25:00 PM	19.60	19.80	0.20
2:26:00 PM	19.20	19.80	0.60
2:27:00 PM	19.00	19.50	0.50
2:28:00 PM	19.50	19.50	0.00
2:29:00 PM	19.80	19.40	-0.40
2:30:00 PM	19.90	19.20	-0.70
2:31:00 PM	20.00	19.20	-0.80
2:32:00 PM	20.00	19.80	-0.20
2:33:00 PM	20.10	20.10	0.00
2:34:00 PM	20.30	20.10	-0.20
2:35:00 PM	20.50	20.20	-0.30
2:36:00 PM	20.60	20.30	-0.30
2:37:00 PM	20.90	20.40	-0.50
2:38:00 PM	21.10	20.50	-0.60
2:39:00 PM	21.20	20.40	-0.80
2:40:00 PM	21.50	20.80	-0.70
2:41:00 PM	21.50	20.80	-0.70
2:42:00 PM	21.60	20.70	-0.90
2:43:00 PM	21.80	21.00	-0.80
2:44:00 PM	21.90	21.20	-0.70
Promedio	20.35	20.21	
Desviación	0.778	0.530	
Max	21.90	21.00	
Min	19.00	19.20	

Certificado No.:

62977

Registrador de Temperatura y Humedad

Fecha de Calibración:


2025-10-30

AM2315

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
2:15:00 PM	50.90	50.80	-0.10
2:16:00 PM	50.50	50.50	0.00
2:17:00 PM	50.50	50.90	0.40
2:18:00 PM	50.10	50.40	0.30
2:19:00 PM	50.00	50.10	0.10
2:20:00 PM	50.00	50.00	0.00
2:21:00 PM	49.80	50.00	0.20
2:22:00 PM	49.70	49.90	0.20
2:23:00 PM	49.50	49.50	0.00
2:24:00 PM	49.40	49.70	0.30
2:25:00 PM	49.20	49.80	0.60
2:26:00 PM	49.10	49.90	0.80
2:27:00 PM	49.00	50.00	1.00
2:28:00 PM	48.80	50.10	1.30
2:29:00 PM	48.50	50.50	2.00
2:30:00 PM	48.90	50.10	1.20
2:31:00 PM	49.50	49.80	0.30
2:32:00 PM	49.80	49.50	-0.30
2:33:00 PM	49.90	49.80	-0.10
2:34:00 PM	50.20	50.10	-0.10
2:35:00 PM	50.30	50.20	-0.10
2:36:00 PM	50.50	50.30	-0.20
2:37:00 PM	50.40	50.40	0.00
2:38:00 PM	50.80	50.80	0.00
2:39:00 PM	50.90	50.90	0.00
2:40:00 PM	51.10	51.00	-0.10
2:41:00 PM	51.50	51.10	-0.40
2:42:00 PM	51.90	51.20	-0.70
2:43:00 PM	51.40	51.30	-0.10
2:44:00 PM	51.80	51.40	-0.40
Promedio	50.13	43.05	
Desviación	0.899	0.544	
Max	51.90	51.30	
Min	48.50	49.50	

Curva de Calibración - TEMPERATURA



Curva de Calibración - % HUMEDAD RELATIVA

