

Certificado No.:

63334

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
9:30:00 AM	15.5	14.7	-0.80
9:31:00 AM	14.9	14.6	-0.30
9:32:00 AM	14.7	14.6	-0.10
9:33:00 AM	14.8	14.5	-0.30
9:34:00 AM	14.6	14.4	-0.20
9:35:00 AM	14.7	14.3	-0.40
9:36:00 AM	14.8	14.2	-0.60
9:37:00 AM	14.9	14.0	-0.90
9:38:00 AM	14.5	14.0	-0.50
9:39:00 AM	14.0	13.8	-0.20
9:40:00 AM	14.0	13.6	-0.40
9:41:00 AM	13.9	13.6	-0.29
9:42:00 AM	13.9	13.3	-0.60
9:43:00 AM	14.0	13.2	-0.80
9:44:00 AM	14.3	13.4	-0.90
9:45:00 AM	14.2	13.7	-0.50
9:46:00 AM	14.4	13.8	-0.60
9:47:00 AM	14.5	13.9	-0.60
9:48:00 AM	14.6	14.0	-0.60
9:49:00 AM	14.7	14.0	-0.70
9:50:00 AM	15.0	14.3	-0.70
9:51:00 AM	15.0	14.2	-0.80
9:52:00 AM	15.2	14.5	-0.70
9:53:00 AM	15.3	14.6	-0.70
9:54:00 AM	15.4	14.7	-0.70
9:55:00 AM	15.5	14.8	-0.70
9:56:00 AM	15.7	14.6	-1.10
9:57:00 AM	15.4	14.5	-0.90
9:58:00 AM	15.0	14.3	-0.70
9:59:00 AM	15.3	14.6	-0.70
Promedio	14.76	14.16	
Desviación	0.520	0.452	
Max	15.70	14.80	
Min	13.90	13.20	

Certificado No.:

63334

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
9:30:00 AM	59.8	60.3	0.50
9:31:00 AM	59.6	60.3	0.70
9:32:00 AM	59.5	59.8	0.30
9:33:00 AM	59.4	59.7	0.30
9:34:00 AM	59.3	59.3	0.00
9:35:00 AM	59.2	59.0	-0.20
9:36:00 AM	59.1	59.0	-0.10
9:37:00 AM	59.0	58.7	-0.30
9:38:00 AM	59.0	58.6	-0.40
9:39:00 AM	59.0	58.4	-0.60
9:40:00 AM	58.9	58.3	-0.60
9:41:00 AM	58.7	58.2	-0.50
9:42:00 AM	58.5	58.0	-0.50
9:43:00 AM	58.4	58.0	-0.40
9:44:00 AM	58.4	57.9	-0.50
9:45:00 AM	58.9	57.8	-1.10
9:46:00 AM	59.0	57.4	-1.60
9:47:00 AM	59.0	57.3	-1.70
9:48:00 AM	59.3	57.2	-2.10
9:49:00 AM	59.4	57.1	-2.30
9:50:00 AM	59.5	57.0	-2.50
9:51:00 AM	59.8	57.0	-2.80
9:52:00 AM	59.9	58.4	-1.50
9:53:00 AM	60.0	58.5	-1.50
9:54:00 AM	60.3	58.6	-1.70
9:55:00 AM	60.4	58.2	-2.20
9:56:00 AM	60.9	58.0	-2.90
9:57:00 AM	61.2	58.0	-3.20
9:58:00 AM	61.5	58.6	-2.90
9:59:00 AM	61.7	58.7	-3.00
Promedio	59.55	58.38	
Desviación	0.872	0.896	
Max	61.70	60.30	
Min	58.40	57.00	

Certificado No.:

63334

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

