



Certificado No.:

63335

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

\*\*\*\*\*

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
10:30:00 AM	14.8	13.9	-0.90
10:31:00 AM	14.7	14.0	-0.70
10:32:00 AM	14.5	14.4	-0.10
10:33:00 AM	14.5	14.4	-0.10
10:34:00 AM	14.5	14.4	-0.10
10:35:00 AM	14.5	14.4	-0.10
10:36:00 AM	14.5	14.2	-0.30
10:37:00 AM	14.2	14.5	0.30
10:38:00 AM	14.1	14.7	0.60
10:39:00 AM	13.8	14.7	0.90
10:40:00 AM	13.7	14.7	1.00
10:41:00 AM	13.6	14.7	1.10
10:42:00 AM	13.6	14.8	1.20
10:43:00 AM	13.6	15.0	1.40
10:44:00 AM	13.6	15.0	1.40
10:45:00 AM	13.4	15.0	1.60
10:46:00 AM	13.2	15.0	1.80
10:47:00 AM	13.0	14.6	1.60
10:48:00 AM	13.0	14.2	1.20
10:49:00 AM	13.0	13.8	0.80
10:50:00 AM	13.0	13.7	0.70
10:51:00 AM	13.0	13.6	0.60
10:52:00 AM	13.8	13.6	-0.20
10:53:00 AM	13.9	13.6	-0.30
10:54:00 AM	14.0	13.5	-0.50
10:55:00 AM	14.4	13.7	-0.70
10:56:00 AM	14.5	13.9	-0.60
10:57:00 AM	14.6	14.7	0.10
10:58:00 AM	15.0	14.8	-0.20
10:59:00 AM	15.0	14.9	-0.10
<b>Promedio</b>	<b>13.97</b>	<b>14.35</b>	
<b>Desviación</b>	<b>0.643</b>	<b>0.499</b>	
<b>Max</b>	<b>15.00</b>	<b>15.00</b>	
<b>Min</b>	<b>13.00</b>	<b>13.50</b>	

Certificado No.:

63335

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

\*\*\*\*\*

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
10:30:00 AM	58.7	55.8	-2.90
10:31:00 AM	58.4	55.9	-2.50
10:32:00 AM	58.3	56.3	-2.00
10:33:00 AM	58.3	56.4	-1.90
10:34:00 AM	58.3	56.7	-1.60
10:35:00 AM	58.3	56.8	-1.50
10:36:00 AM	58.3	57.3	-1.00
10:37:00 AM	59.0	57.5	-1.50
10:38:00 AM	58.7	57.8	-0.90
10:39:00 AM	58.9	57.4	-1.50
10:40:00 AM	59.0	57.9	-1.10
10:41:00 AM	59.0	58.3	-0.70
10:42:00 AM	59.0	58.4	-0.60
10:43:00 AM	59.0	58.7	-0.30
10:44:00 AM	59.0	58.9	-0.10
10:45:00 AM	59.0	58.6	-0.40
10:46:00 AM	59.3	58.4	-0.90
10:47:00 AM	59.4	58.3	-1.10
10:48:00 AM	59.5	58.0	-1.50
10:49:00 AM	59.5	58.0	-1.50
10:50:00 AM	59.5	57.9	-1.60
10:51:00 AM	59.5	57.8	-1.70
10:52:00 AM	59.5	57.7	-1.80
10:53:00 AM	59.4	57.6	-1.80
10:54:00 AM	59.3	57.8	-1.50
10:55:00 AM	59.2	57.9	-1.30
10:56:00 AM	59.0	58.2	-0.80
10:57:00 AM	58.7	58.4	-0.30
10:58:00 AM	58.3	58.4	0.10
10:59:00 AM	58.2	57.9	-0.30
<b>Promedio</b>	<b>58.92</b>	<b>57.70</b>	
<b>Desviación</b>	<b>0.447</b>	<b>0.813</b>	
<b>Max</b>	<b>59.50</b>	<b>58.90</b>	
<b>Min</b>	<b>58.20</b>	<b>55.80</b>	

Certificado No.:

63335

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

\*\*\*\*\*

