

Certificado No.:

63336

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

HWA-SUNG-THERMO

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
11:10:00 AM	15.8	13.7	-2.10
11:11:00 AM	15.4	13.8	-1.60
11:12:00 AM	15.3	13.3	-2.00
11:13:00 AM	15.3	13.1	-2.20
11:14:00 AM	15.2	13.8	-1.40
11:15:00 AM	15.1	13.9	-1.20
11:16:00 AM	14.9	14.3	-0.60
11:17:00 AM	14.8	14.2	-0.60
11:18:00 AM	14.7	14.5	-0.20
11:19:00 AM	14.6	14.8	0.20
11:20:00 AM	14.3	14.9	0.60
11:21:00 AM	13.9	15.0	1.10
11:22:00 AM	13.8	15.0	1.20
11:23:00 AM	13.7	15.7	2.00
11:24:00 AM	13.6	15.9	2.30
11:25:00 AM	13.6	15.9	2.30
11:26:00 AM	13.6	16.1	2.50
11:27:00 AM	13.6	16.3	2.70
11:28:00 AM	13.8	16.2	2.40
11:29:00 AM	13.9	15.4	1.50
11:30:00 AM	14.3	14.9	0.60
11:31:00 AM	14.3	14.8	0.50
11:32:00 AM	14.6	14.5	-0.10
11:33:00 AM	14.7	14.3	-0.40
11:34:00 AM	14.8	14.3	-0.50
11:35:00 AM	15.0	14.5	-0.50
11:36:00 AM	15.4	14.6	-0.80
11:37:00 AM	15.5	14.8	-0.70
11:38:00 AM	15.8	15.0	-0.80
11:39:00 AM	15.9	15.0	-0.90
Promedio	14.64	14.75	
Desviación	0.740	0.834	
Max	15.90	16.30	
Min	13.60	13.10	

Certificado No.:

63336

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

HWA-SUNG-THERMO

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
11:10:00 AM	60.3	58.5	-1.80
11:11:00 AM	60.4	58.4	-2.00
11:12:00 AM	60.8	58.3	-2.50
11:13:00 AM	60.9	58.2	-2.70
11:14:00 AM	60.4	58.0	-2.40
11:15:00 AM	60.3	58.0	-2.30
11:16:00 AM	60.1	59.2	-0.90
11:17:00 AM	60.0	59.3	-0.70
11:18:00 AM	59.5	59.4	-0.10
11:19:00 AM	59.4	59.4	0.00
11:20:00 AM	59.3	59.1	-0.20
11:21:00 AM	59.2	59.0	-0.20
11:22:00 AM	59.1	58.5	-0.60
11:23:00 AM	59.3	58.4	-0.90
11:24:00 AM	59.4	58.3	-1.10
11:25:00 AM	59.5	58.2	-1.30
11:26:00 AM	58.9	58.0	-0.90
11:27:00 AM	58.7	58.0	-0.70
11:28:00 AM	58.6	57.3	-1.30
11:29:00 AM	58.9	57.4	-1.50
11:30:00 AM	59.0	57.5	-1.50
11:31:00 AM	59.0	57.3	-1.70
11:32:00 AM	59.3	57.2	-2.10
11:33:00 AM	59.4	57.0	-2.40
11:34:00 AM	59.8	57.4	-2.40
11:35:00 AM	60.3	57.8	-2.50
11:36:00 AM	60.4	57.9	-2.50
11:37:00 AM	60.6	58.3	-2.30
11:38:00 AM	60.7	58.5	-2.20
11:39:00 AM	60.9	58.9	-2.00
Promedio	59.75	58.22	
Desviación	0.708	0.686	
Max	60.90	59.40	
Min	58.60	57.00	

Certificado No.:

63336

Registrador de Temperatura y Humedad

Fecha de Calibración:

2025-11-15

HWA-SUNG-THERMO

