



Certificado No.:

64696

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-01-06

\*\*\*\*\*

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
1:30:00 PM	18.9	19.1	0.20
1:31:00 PM	18.5	19.0	0.50
1:32:00 PM	18.4	18.5	0.10
1:33:00 PM	18.2	18.4	0.20
1:34:00 PM	18.0	18.2	0.20
1:35:00 PM	18.0	18.1	0.10
1:36:00 PM	17.9	18.0	0.10
1:37:00 PM	17.5	17.8	0.30
1:38:00 PM	17.4	17.4	0.00
1:39:00 PM	17.2	17.1	-0.10
1:40:00 PM	17.0	17.0	0.00
1:41:00 PM	17.0	17.0	0.00
1:42:00 PM	16.8	16.8	0.00
1:43:00 PM	16.7	16.8	0.10
1:44:00 PM	16.5	16.4	-0.10
1:45:00 PM	16.0	16.2	0.20
1:46:00 PM	16.0	16.1	0.10
1:47:00 PM	15.8	16.0	0.20
1:48:00 PM	15.9	15.8	-0.10
1:49:00 PM	15.3	15.9	0.60
1:50:00 PM	16.1	15.7	-0.40
1:51:00 PM	16.4	15.8	-0.60
1:52:00 PM	16.5	15.9	-0.60
1:53:00 PM	16.8	16.0	-0.80
1:54:00 PM	16.9	16.1	-0.80
1:55:00 PM	17.0	16.4	-0.60
1:56:00 PM	17.4	16.9	-0.50
1:57:00 PM	17.5	16.7	-0.80
1:58:00 PM	17.8	17.5	-0.30
1:59:00 PM	17.9	17.2	-0.70
<b>Promedio</b>	<b>17.11</b>	<b>16.99</b>	
<b>Desviación</b>	<b>0.898</b>	<b>1.009</b>	
<b>Max</b>	<b>18.90</b>	<b>19.10</b>	
<b>Min</b>	<b>15.30</b>	<b>15.70</b>	

Certificado No.:

64696

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-01-06

\*\*\*\*\*

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
1:30:00 PM	64.8	63.8	-1.00
1:31:00 PM	64.7	63.7	-1.00
1:32:00 PM	64.5	63.5	-1.00
1:33:00 PM	64.2	63.4	-0.80
1:34:00 PM	64.1	63.1	-1.00
1:35:00 PM	64.0	63.0	-1.00
1:36:00 PM	63.8	62.8	-1.00
1:37:00 PM	63.5	62.4	-1.10
1:38:00 PM	63.4	62.7	-0.70
1:39:00 PM	63.1	62.1	-1.00
1:40:00 PM	63.0	61.9	-1.10
1:41:00 PM	62.8	61.8	-1.00
1:42:00 PM	62.9	61.5	-1.40
1:43:00 PM	62.4	61.4	-1.00
1:44:00 PM	62.1	61.2	-0.90
1:45:00 PM	62.0	61.1	-0.90
1:46:00 PM	62.0	61.0	-1.00
1:47:00 PM	61.8	61.0	-0.80
1:48:00 PM	61.7	62.4	0.70
1:49:00 PM	61.5	62.8	1.30
1:50:00 PM	61.2	62.9	1.70
1:51:00 PM	62.5	63.1	0.60
1:52:00 PM	62.8	63.4	0.60
1:53:00 PM	62.9	63.7	0.80
1:54:00 PM	62.7	63.4	0.70
1:55:00 PM	62.5	63.1	0.60
1:56:00 PM	62.4	63.2	0.80
1:57:00 PM	62.4	63.0	0.60
1:58:00 PM	62.1	62.9	0.80
1:59:00 PM	63.5	62.9	-0.60
<b>Promedio</b>	<b>62.91</b>	<b>62.61</b>	
<b>Desviación</b>	<b>0.967</b>	<b>0.867</b>	
<b>Max</b>	<b>64.80</b>	<b>63.80</b>	
<b>Min</b>	<b>61.20</b>	<b>61.00</b>	

Certificado No.:

64696

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-01-06

\*\*\*\*\*

