



Certificado No.:

65550

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-01-27

\*\*\*\*\*

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
2:30:00 PM	21.1	21.0	-0.10
2:31:00 PM	21.0	21.0	0.00
2:32:00 PM	21.0	21.0	0.00
2:33:00 PM	21.0	20.5	-0.50
2:34:00 PM	20.9	20.1	-0.80
2:35:00 PM	20.8	20.5	-0.30
2:36:00 PM	20.5	20.6	0.10
2:37:00 PM	20.7	20.4	-0.30
2:38:00 PM	20.6	20.8	0.20
2:39:00 PM	20.2	21.1	0.90
2:40:00 PM	20.0	21.0	1.00
2:41:00 PM	20.0	20.5	0.50
2:42:00 PM	20.0	20.1	0.10
2:43:00 PM	19.8	20.0	0.20
2:44:00 PM	19.5	20.0	0.50
2:45:00 PM	19.9	19.8	-0.10
2:46:00 PM	20.0	19.5	-0.50
2:47:00 PM	20.0	19.9	-0.10
2:48:00 PM	20.5	19.5	-1.00
2:49:00 PM	20.8	19.9	-0.90
2:50:00 PM	20.9	20.1	-0.80
2:51:00 PM	21.1	20.5	-0.60
2:52:00 PM	21.5	21.1	-0.40
2:53:00 PM	21.8	21.5	-0.30
2:54:00 PM	22.0	21.8	-0.20
2:55:00 PM	22.0	22.0	0.00
2:56:00 PM	22.1	22.1	0.00
2:57:00 PM	22.4	22.5	0.10
2:58:00 PM	22.4	22.4	0.00
2:59:00 PM	22.6	22.5	-0.10
<b>Promedio</b>	<b>20.90</b>	<b>20.79</b>	
<b>Desviación</b>	<b>0.865</b>	<b>0.883</b>	
<b>Max</b>	<b>22.60</b>	<b>22.50</b>	
<b>Min</b>	<b>19.50</b>	<b>19.50</b>	

Certificado No.:

65550

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-01-27

\*\*\*\*\*

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
2:30:00 PM	64.2	63.3	-0.90
2:31:00 PM	64.1	63.1	-1.00
2:32:00 PM	64.0	62.5	-1.50
2:33:00 PM	64.0	62.8	-1.20
2:34:00 PM	63.8	62.9	-0.90
2:35:00 PM	63.9	62.4	-1.50
2:36:00 PM	63.5	62.7	-0.80
2:37:00 PM	63.2	62.5	-0.70
2:38:00 PM	63.1	63.0	-0.10
2:39:00 PM	63.0	63.0	0.00
2:40:00 PM	63.0	63.1	0.10
2:41:00 PM	62.9	63.2	0.30
2:42:00 PM	62.8	63.5	0.70
2:43:00 PM	62.7	63.5	0.80
2:44:00 PM	62.9	63.1	0.21
2:45:00 PM	62.9	63.2	0.30
2:46:00 PM	63.0	63.0	0.00
2:47:00 PM	63.0	63.0	0.00
2:48:00 PM	63.4	63.8	0.40
2:49:00 PM	63.2	63.9	0.70
2:50:00 PM	63.5	63.9	0.40
2:51:00 PM	63.9	63.8	-0.10
2:52:00 PM	64.0	63.3	-0.70
2:53:00 PM	64.1	63.2	-0.90
2:54:00 PM	64.2	63.1	-1.10
2:55:00 PM	64.2	63.0	-1.20
2:56:00 PM	64.8	62.8	-2.00
2:57:00 PM	64.9	62.5	-2.40
2:58:00 PM	65.1	62.4	-2.70
2:59:00 PM	65.5	62.3	-3.20
<b>Promedio</b>	<b>63.69</b>	<b>63.06</b>	
<b>Desviación</b>	<b>0.743</b>	<b>0.446</b>	
<b>Max</b>	<b>65.50</b>	<b>63.90</b>	
<b>Min</b>	<b>62.70</b>	<b>62.40</b>	

Certificado No.:

65550


Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-01-27

\*\*\*\*\*

### Curva de Calibración - TEMPERATURA



### Curva de Calibración - % HUMEDAD RELATIVA

