

Certificado No.:

65591

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-01-28

TRACKLINK

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
10:30:00 AM	19.1	19.2	0.10
10:31:00 AM	19.2	19.2	0.00
10:32:00 AM	19.3	19.2	-0.10
10:33:00 AM	19.3	19.2	-0.10
10:34:00 AM	19.3	19.2	-0.10
10:35:00 AM	19.4	19.5	0.10
10:36:00 AM	19.5	19.5	0.00
10:37:00 AM	19.5	19.5	0.00
10:38:00 AM	19.6	19.5	-0.10
10:39:00 AM	19.6	19.5	-0.10
10:40:00 AM	19.7	19.7	0.00
10:41:00 AM	19.7	19.7	0.00
10:42:00 AM	19.8	19.7	-0.10
10:43:00 AM	19.9	19.7	-0.20
10:44:00 AM	19.9	19.7	-0.20
10:45:00 AM	19.9	20.0	0.10
10:46:00 AM	20.0	20.0	0.00
10:47:00 AM	20.0	20.0	0.00
10:48:00 AM	20.1	20.0	-0.10
10:49:00 AM	20.1	20.0	-0.10
10:50:00 AM	20.2	20.2	0.00
10:51:00 AM	20.2	20.2	0.00
10:52:00 AM	20.2	20.2	0.00
10:53:00 AM	20.3	20.2	-0.10
10:54:00 AM	20.3	20.2	-0.10
10:55:00 AM	20.4	20.5	0.10
10:56:00 AM	20.4	20.5	0.10
10:57:00 AM	20.5	20.5	0.00
10:58:00 AM	20.5	20.5	0.00
10:59:00 AM	20.6	20.5	-0.10
Promedio	19.88	19.9	
Desviación	0.433	0.442	
Max	20.60	20.50	
Min	19.10	19.20	

Certificado No.:

65591

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-01-28

TRACKLINK

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
10:30:00 AM	64.8	71.9	7.10
10:31:00 AM	64.8	71.9	7.10
10:32:00 AM	64.8	71.9	7.10
10:33:00 AM	64.8	71.9	7.10
10:34:00 AM	65.3	71.9	6.60
10:35:00 AM	64.8	70.9	6.10
10:36:00 AM	64.8	70.9	6.10
10:37:00 AM	65.3	70.9	5.60
10:38:00 AM	64.8	70.9	6.10
10:39:00 AM	64.8	70.9	6.10
10:40:00 AM	65.3	70.4	5.10
10:41:00 AM	65.3	70.4	5.10
10:42:00 AM	65.3	70.4	5.10
10:43:00 AM	65.3	70.4	5.10
10:44:00 AM	65.3	70.4	5.10
10:45:00 AM	65.3	69.4	4.10
10:46:00 AM	65.3	69.4	4.10
10:47:00 AM	65.3	69.4	4.10
10:48:00 AM	65.4	69.4	4.00
10:49:00 AM	64.9	69.4	4.50
10:50:00 AM	64.9	68.4	3.50
10:51:00 AM	64.9	68.4	3.50
10:52:00 AM	64.9	68.4	3.50
10:53:00 AM	64.9	68.4	3.50
10:54:00 AM	64.9	68.4	3.50
10:55:00 AM	64.9	67.4	2.50
10:56:00 AM	64.9	67.4	2.50
10:57:00 AM	64.9	67.4	2.50
10:58:00 AM	64.9	67.4	2.50
10:59:00 AM	64.9	67.4	2.50
Promedio	65.02	69.7	
Desviación	0.225	1.544	
Max	65.40	71.90	
Min	64.80	67.40	

Certificado No.:

65591


Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-01-28

TRACKLINK

Curva de Calibración - TEMPERATURA



Curva de Calibración - % HUMEDAD RELATIVA

