



Certificado No.:

66100

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-02-11

MAXIM

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
10:30:00 AM	20.0	19.7	-0.30
10:31:00 AM	20.0	19.7	-0.30
10:32:00 AM	20.1	20.0	-0.10
10:33:00 AM	20.1	20.0	-0.10
10:34:00 AM	20.2	20.2	0.00
10:35:00 AM	20.3	20.2	-0.10
10:36:00 AM	20.3	20.5	0.20
10:37:00 AM	20.3	20.5	0.20
10:38:00 AM	20.4	20.7	0.30
10:39:00 AM	20.5	20.7	0.20
10:40:00 AM	20.5	20.7	0.20
10:41:00 AM	20.5	20.7	0.20
10:42:00 AM	20.5	21.0	0.50
10:43:00 AM	20.5	21.0	0.50
10:44:00 AM	20.7	21.2	0.50
10:45:00 AM	20.7	21.2	0.50
10:46:00 AM	20.9	21.2	0.30
10:47:00 AM	20.9	21.5	0.60
10:48:00 AM	21.1	21.5	0.40
10:49:00 AM	21.1	21.5	0.40
10:50:00 AM	21.4	21.7	0.30
10:51:00 AM	21.5	21.7	0.20
10:52:00 AM	21.6	21.7	0.10
10:53:00 AM	21.6	21.7	0.10
10:54:00 AM	21.7	21.7	0.00
10:55:00 AM	21.9	21.7	-0.20
10:56:00 AM	21.9	21.7	-0.20
10:57:00 AM	22.0	22.2	0.20
10:58:00 AM	22.0	22.2	0.20
10:59:00 AM	22.1	22.5	0.40
<b>Promedio</b>	<b>20.91</b>	<b>21.08</b>	
<b>Desviación</b>	<b>0.689</b>	<b>0.760</b>	
<b>Max</b>	<b>22.10</b>	<b>22.20</b>	
<b>Min</b>	<b>20.00</b>	<b>19.70</b>	

Certificado No.:

66100

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-02-11

MAXIM

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
10:30:00 AM	40.21	39.54	-0.67
10:31:00 AM	40.32	40.48	0.16
10:32:00 AM	40.35	40.41	0.06
10:33:00 AM	40.51	40.11	-0.40
10:34:00 AM	40.55	39.91	-0.64
10:35:00 AM	40.66	39.58	-1.08
10:36:00 AM	40.74	38.87	-1.87
10:37:00 AM	40.79	38.56	-2.23
10:38:00 AM	40.43	38.25	-2.18
10:39:00 AM	40.11	38.11	-2.00
10:40:00 AM	40.01	38.88	-1.13
10:41:00 AM	39.78	38.94	-0.84
10:42:00 AM	39.64	39.05	-0.59
10:43:00 AM	39.92	39.26	-0.66
10:44:00 AM	40.12	39.44	-0.68
10:45:00 AM	40.41	39.50	-0.91
10:46:00 AM	40.55	39.78	-0.77
10:47:00 AM	40.84	39.91	-0.93
10:48:00 AM	40.99	40.11	-0.88
10:49:00 AM	41.21	40.34	-0.87
10:50:00 AM	41.37	40.25	-1.12
10:51:00 AM	41.95	40.48	-1.47
10:52:00 AM	42.34	40.69	-1.65
10:53:00 AM	42.75	40.83	-1.92
10:54:00 AM	42.97	41.38	-1.59
10:55:00 AM	43.46	41.94	-1.52
10:56:00 AM	43.91	42.56	-1.35
10:57:00 AM	44.16	42.77	-1.39
10:58:00 AM	44.73	43.01	-1.72
10:59:00 AM	45.07	43.47	-1.60
<b>Promedio</b>	<b>41.36</b>	<b>40.21</b>	
<b>Desviación</b>	<b>1.56</b>	<b>1.395</b>	
<b>Max</b>	<b>45.07</b>	<b>43.01</b>	
<b>Min</b>	<b>39.64</b>	<b>38.11</b>	

Certificado No.:

66100

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-02-11

MAXIM

