

Certificado No.:

66894

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-06

SMART

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
3:15:00 PM	29.29	26.30	-2.99
3:16:00 PM	29.05	26.30	-2.75
3:17:00 PM	28.86	26.40	-2.46
3:18:00 PM	28.85	26.50	-2.35
3:19:00 PM	28.71	26.40	-2.31
3:20:00 PM	28.64	26.50	-2.14
3:21:00 PM	28.33	26.70	-1.63
3:22:00 PM	28.27	26.70	-1.57
3:23:00 PM	28.12	26.70	-1.42
3:24:00 PM	28.00	26.80	-1.20
3:25:00 PM	27.87	26.90	-0.97
3:26:00 PM	27.75	26.90	-0.85
3:27:00 PM	27.61	27.00	-0.61
3:28:00 PM	27.46	27.00	-0.46
3:29:00 PM	27.36	27.00	-0.36
3:30:00 PM	27.28	27.10	-0.18
3:31:00 PM	27.12	27.00	-0.12
3:32:00 PM	27.11	27.10	-0.01
3:33:00 PM	26.98	27.20	0.22
3:34:00 PM	26.83	27.30	0.47
3:35:00 PM	26.71	27.30	0.59
3:36:00 PM	26.61	27.40	0.79
3:37:00 PM	26.43	27.40	0.97
3:38:00 PM	26.41	27.40	0.99
3:39:00 PM	26.21	27.50	1.29
3:40:00 PM	26.15	27.60	1.45
3:41:00 PM	26.07	27.60	1.53
3:42:00 PM	25.99	27.70	1.71
3:43:00 PM	25.85	27.70	1.85
3:44:00 PM	25.76	27.80	2.04
Promedio	27.39	27.04	
Desviación	1.054	0.447	
Max	29.29	27.70	
Min	25.76	26.30	

Certificado No.:

66894

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-06

SMART

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
3:15:00 PM	50.27	44.90	-5.37
3:16:00 PM	50.37	45.30	-5.07
3:17:00 PM	49.97	45.90	-4.07
3:18:00 PM	49.87	46.70	-3.17
3:19:00 PM	49.77	45.80	-3.97
3:20:00 PM	49.27	46.20	-3.07
3:21:00 PM	49.57	46.00	-3.57
3:22:00 PM	49.27	43.80	-5.47
3:23:00 PM	49.17	45.80	-3.37
3:24:00 PM	48.97	46.70	-2.27
3:25:00 PM	48.77	46.50	-2.27
3:26:00 PM	48.77	46.30	-2.47
3:27:00 PM	48.47	46.00	-2.47
3:28:00 PM	48.27	45.80	-2.47
3:29:00 PM	47.97	45.80	-2.17
3:30:00 PM	47.77	44.00	-3.77
3:31:00 PM	47.57	44.80	-2.77
3:32:00 PM	47.67	45.20	-2.47
3:33:00 PM	47.27	45.80	-1.47
3:34:00 PM	47.17	45.10	-2.07
3:35:00 PM	46.87	45.70	-1.17
3:36:00 PM	46.57	45.10	-1.47
3:37:00 PM	46.67	45.20	-1.47
3:38:00 PM	46.27	45.00	-1.27
3:39:00 PM	46.17	44.00	-2.17
3:40:00 PM	45.97	43.30	-2.67
3:41:00 PM	45.87	43.70	-2.17
3:42:00 PM	45.77	42.50	-3.27
3:43:00 PM	45.57	44.20	-1.37
3:44:00 PM	45.87	43.20	-2.67
Promedio	47.93	45.14	
Desviación	1.528	1.105	
Max	50.37	46.70	
Min	45.57	42.50	

Certificado No.:

66894

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-06

SMART

