

Certificado No.:

66900

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-02-27

FULL GAUGE

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
12:50:00 PM	22.70	22.70	0.00
12:51:00 PM	22.70	22.70	0.00
12:52:00 PM	22.70	22.60	-0.10
12:53:00 PM	22.70	22.70	0.00
12:54:00 PM	22.70	22.60	-0.10
12:55:00 PM	22.80	22.60	-0.20
12:56:00 PM	22.70	22.70	0.00
12:57:00 PM	22.70	22.70	0.00
12:58:00 PM	22.70	22.70	0.00
12:59:00 PM	22.80	22.70	-0.10
1:00:00 PM	22.70	22.70	0.00
1:01:00 PM	22.70	22.70	0.00
1:02:00 PM	22.70	22.70	0.00
1:03:00 PM	22.70	22.70	0.00
1:04:00 PM	22.70	22.70	0.00
1:05:00 PM	22.70	22.70	0.00
1:06:00 PM	22.60	22.70	0.10
1:07:00 PM	22.60	22.70	0.10
1:08:00 PM	22.60	22.70	0.10
1:09:00 PM	22.60	22.70	0.10
1:10:00 PM	22.60	22.70	0.10
1:11:00 PM	22.60	22.70	0.10
1:12:00 PM	22.70	22.70	0.00
1:13:00 PM	22.60	22.70	0.10
1:14:00 PM	22.70	22.70	0.00
1:15:00 PM	22.70	22.70	0.00
1:16:00 PM	22.60	22.70	0.10
1:17:00 PM	22.60	22.70	0.10
1:18:00 PM	22.60	22.70	0.10
1:19:00 PM	22.60	22.70	0.10
Promedio	22.67	22.69	
Desviación	0.060	0.031	
Max	22.80	22.70	
Min	22.60	22.60	

Certificado No.:

66900

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-02-27

FULL GAUGE

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
12:50:00 PM	55.60	53.50	-2.10
12:51:00 PM	55.60	53.70	-1.90
12:52:00 PM	55.60	53.80	-1.80
12:53:00 PM	55.20	53.60	-1.60
12:54:00 PM	55.20	53.30	-1.90
12:55:00 PM	55.20	53.50	-1.70
12:56:00 PM	55.20	53.30	-1.90
12:57:00 PM	55.20	53.60	-1.60
12:58:00 PM	55.20	53.60	-1.60
12:59:00 PM	55.20	53.80	-1.40
1:00:00 PM	54.70	53.90	-0.80
1:01:00 PM	54.70	53.80	-0.90
1:02:00 PM	54.70	53.80	-0.90
1:03:00 PM	54.70	53.80	-0.90
1:04:00 PM	54.70	53.80	-0.90
1:05:00 PM	54.70	53.70	-1.00
1:06:00 PM	54.70	53.60	-1.10
1:07:00 PM	54.70	53.70	-1.00
1:08:00 PM	54.20	53.60	-0.60
1:09:00 PM	54.20	53.70	-0.50
1:10:00 PM	54.20	53.70	-0.50
1:11:00 PM	54.20	53.70	-0.50
1:12:00 PM	54.20	53.80	-0.40
1:13:00 PM	54.20	53.80	-0.40
1:14:00 PM	54.20	53.70	-0.50
1:15:00 PM	54.20	53.70	-0.50
1:16:00 PM	54.20	53.70	-0.50
1:17:00 PM	54.20	53.70	-0.50
1:18:00 PM	54.20	53.80	-0.40
1:19:00 PM	54.20	53.80	-0.40
Promedio	54.71	53.68	
Desviación	0.498	0.142	
Max	55.60	53.90	
Min	54.20	53.30	

