

Certificado No.:

66910

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-06

SMART

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
3:15:00 PM	25.19	23.50	-1.69
3:16:00 PM	24.95	23.50	-1.45
3:17:00 PM	24.76	23.50	-1.26
3:18:00 PM	24.75	23.50	-1.25
3:19:00 PM	24.61	23.50	-1.11
3:20:00 PM	24.54	23.50	-1.04
3:21:00 PM	24.23	23.50	-0.73
3:22:00 PM	24.17	23.50	-0.67
3:23:00 PM	24.02	23.50	-0.52
3:24:00 PM	23.90	23.50	-0.40
3:25:00 PM	23.77	24.00	0.23
3:26:00 PM	23.65	24.00	0.35
3:27:00 PM	23.51	24.00	0.49
3:28:00 PM	23.36	24.10	0.74
3:29:00 PM	23.26	24.10	0.84
3:30:00 PM	23.18	24.00	0.82
3:31:00 PM	23.02	24.00	0.98
3:32:00 PM	23.01	24.00	0.99
3:33:00 PM	22.88	23.70	0.82
3:34:00 PM	22.73	23.70	0.97
3:35:00 PM	22.61	23.70	1.09
3:36:00 PM	22.51	23.70	1.19
3:37:00 PM	22.33	23.70	1.37
3:38:00 PM	22.31	23.70	1.39
3:39:00 PM	22.11	23.70	1.59
3:40:00 PM	22.05	23.70	1.65
3:41:00 PM	21.97	23.70	1.73
3:42:00 PM	21.89	23.70	1.81
3:43:00 PM	21.75	23.20	1.45
3:44:00 PM	21.66	23.20	1.54
Promedio	23.29	23.69	
Desviación	1.054	0.246	
Max	25.19	24.10	
Min	21.66	23.20	

Certificado No.:

66910

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-06

SMART

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
3:15:00 PM	53.77	48.00	-5.77
3:16:00 PM	53.87	48.00	-5.87
3:17:00 PM	53.47	48.00	-5.47
3:18:00 PM	53.37	48.00	-5.37
3:19:00 PM	53.27	48.00	-5.27
3:20:00 PM	52.77	48.00	-4.77
3:21:00 PM	53.07	48.00	-5.07
3:22:00 PM	52.77	48.00	-4.77
3:23:00 PM	52.67	48.00	-4.67
3:24:00 PM	52.47	48.00	-4.47
3:25:00 PM	52.27	48.00	-4.27
3:26:00 PM	52.27	48.00	-4.27
3:27:00 PM	51.97	47.60	-4.37
3:28:00 PM	51.77	47.60	-4.17
3:29:00 PM	51.47	47.60	-3.87
3:30:00 PM	51.27	47.60	-3.67
3:31:00 PM	51.07	47.60	-3.47
3:32:00 PM	51.17	47.80	-3.37
3:33:00 PM	50.77	47.80	-2.97
3:34:00 PM	50.67	47.80	-2.87
3:35:00 PM	50.37	47.80	-2.57
3:36:00 PM	50.07	47.90	-2.17
3:37:00 PM	50.17	47.90	-2.27
3:38:00 PM	49.77	47.90	-1.87
3:39:00 PM	49.67	47.90	-1.77
3:40:00 PM	49.47	47.90	-1.57
3:41:00 PM	49.37	47.90	-1.47
3:42:00 PM	49.27	47.90	-1.37
3:43:00 PM	49.07	47.90	-1.17
3:44:00 PM	49.37	49.40	0.03
Promedio	51.43	47.93	
Desviación	1.528	0.313	
Max	53.87	48.00	
Min	49.07	47.60	

Certificado No.:

66910

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-06

SMART

