

Certificado No.:

66948

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-10

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
8:03:00 AM	18.5	18.9	0.40
8:04:00 AM	18.5	18.8	0.30
8:05:00 AM	18.4	18.7	0.30
8:06:00 AM	18.2	18.7	0.50
8:07:00 AM	18.1	18.6	0.50
8:08:00 AM	17.5	18.1	0.60
8:09:00 AM	17.6	17.0	-0.60
8:10:00 AM	17.2	17.4	0.20
8:11:00 AM	17.4	17.5	0.10
8:12:00 AM	17.8	17.6	-0.20
8:13:00 AM	17.9	17.8	-0.10
8:14:00 AM	18.2	17.9	-0.30
8:15:00 AM	18.3	18.0	-0.30
8:16:00 AM	18.5	18.4	-0.10
8:17:00 AM	18.8	18.7	-0.10
8:18:00 AM	18.9	18.9	0.00
8:19:00 AM	19.7	19.0	-0.70
8:20:00 AM	19.8	19.5	-0.30
8:21:00 AM	20.0	19.7	-0.30
8:22:00 AM	20.1	19.8	-0.30
8:23:00 AM	20.4	19.9	-0.50
8:24:00 AM	20.5	20.0	-0.50
8:25:00 AM	19.5	20.0	0.50
8:26:00 AM	19.2	19.8	0.60
8:27:00 AM	19.1	19.5	0.40
8:28:00 AM	18.7	19.1	0.40
8:29:00 AM	18.4	18.4	0.00
8:30:00 AM	18.2	18.2	0.00
8:31:00 AM	17.5	18.2	0.70
8:32:00 AM	17.1	18.0	0.90
Promedio	18.60	18.67	
Desviación	0.954	0.842	
Max	20.50	20.00	
Min	17.10	17.00	

Certificado No.:

66948

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-10

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
8:03:00 AM	69.8	66.5	-3.30
8:04:00 AM	69.7	66.8	-2.90
8:05:00 AM	69.4	66.9	-2.50
8:06:00 AM	69.5	67.2	-2.30
8:07:00 AM	69.2	67.2	-2.00
8:08:00 AM	68.8	66.4	-2.40
8:09:00 AM	68.4	66.5	-1.90
8:10:00 AM	68.2	66.1	-2.10
8:11:00 AM	68.1	66.0	-2.10
8:12:00 AM	68.0	67.1	-0.90
8:13:00 AM	67.5	67.2	-0.30
8:14:00 AM	67.8	67.2	-0.60
8:15:00 AM	67.9	67.3	-0.60
8:16:00 AM	68.4	66.8	-1.60
8:17:00 AM	67.9	66.5	-1.40
8:18:00 AM	67.4	66.8	-0.60
8:19:00 AM	66.8	66.1	-0.70
8:20:00 AM	66.4	65.4	-1.00
8:21:00 AM	66.1	65.2	-0.90
8:22:00 AM	66.0	65.8	-0.20
8:23:00 AM	65.4	65.7	0.30
8:24:00 AM	65.2	65.1	-0.10
8:25:00 AM	65.3	64.8	-0.50
8:26:00 AM	65.8	64.7	-1.10
8:27:00 AM	64.8	64.2	-0.60
8:28:00 AM	64.5	64.0	-0.50
8:29:00 AM	64.8	64.0	-0.80
8:30:00 AM	64.9	63.4	-1.50
8:31:00 AM	65.1	63.5	-1.60
8:32:00 AM	65.8	63.8	-2.00
Promedio	67.10	65.81	
Desviación	1.691	1.251	
Max	69.80	67.30	
Min	64.50	63.40	

Certificado No.:

66948

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-10

