



Certificado No.:

66949

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-10

\*\*\*\*\*

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
8:10:00 AM	18.5	19.1	0.60
8:11:00 AM	18.4	18.8	0.40
8:12:00 AM	18.2	18.7	0.50
8:13:00 AM	18.1	18.9	0.80
8:14:00 AM	18.0	18.5	0.50
8:15:00 AM	18.0	18.4	0.40
8:16:00 AM	18.0	18.1	0.10
8:17:00 AM	17.5	18.0	0.50
8:18:00 AM	17.4	18.0	0.60
8:19:00 AM	17.2	17.9	0.70
8:20:00 AM	17.0	17.8	0.80
8:21:00 AM	16.8	17.4	0.60
8:22:00 AM	16.4	17.1	0.70
8:23:00 AM	16.5	17.0	0.50
8:24:00 AM	16.2	17.0	0.80
8:25:00 AM	16.0	16.9	0.90
8:26:00 AM	15.8	16.2	0.40
8:27:00 AM	15.9	16.7	0.80
8:28:00 AM	15.4	16.1	0.70
8:29:00 AM	15.8	16.0	0.20
8:30:00 AM	16.4	16.4	0.00
8:31:00 AM	16.8	16.8	0.00
8:32:00 AM	16.9	16.9	0.00
8:33:00 AM	17.4	16.7	-0.70
8:34:00 AM	17.5	17.0	-0.50
8:35:00 AM	17.9	17.5	-0.40
8:36:00 AM	18.0	17.8	-0.20
8:37:00 AM	18.4	17.9	-0.50
8:38:00 AM	18.5	18.0	-0.50
8:39:00 AM	18.9	18.1	-0.80
<b>Promedio</b>	<b>17.26</b>	<b>17.52</b>	
<b>Desviación</b>	<b>0.977</b>	<b>0.870</b>	
<b>Max</b>	<b>18.90</b>	<b>19.10</b>	
<b>Min</b>	<b>15.40</b>	<b>16.00</b>	

Certificado No.:

66949

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-10

\*\*\*\*\*

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
8:10:00 AM	59.8	56.8	-3.00
8:11:00 AM	59.4	56.4	-3.00
8:12:00 AM	59.1	56.8	-2.30
8:13:00 AM	59.0	56.9	-2.10
8:14:00 AM	58.4	56.8	-1.60
8:15:00 AM	58.6	56.4	-2.20
8:16:00 AM	58.2	56.8	-1.40
8:17:00 AM	58.1	56.1	-2.00
8:18:00 AM	58.0	56.0	-2.00
8:19:00 AM	58.0	55.8	-2.20
8:20:00 AM	57.4	55.7	-1.70
8:21:00 AM	57.1	55.4	-1.70
8:22:00 AM	57.0	55.1	-1.90
8:23:00 AM	56.8	55.0	-1.80
8:24:00 AM	56.9	55.0	-1.90
8:25:00 AM	56.4	54.8	-1.60
8:26:00 AM	56.1	54.7	-1.40
8:27:00 AM	56.0	54.5	-1.50
8:28:00 AM	55.7	54.0	-1.70
8:29:00 AM	55.4	53.8	-1.60
8:30:00 AM	55.1	53.7	-1.40
8:31:00 AM	55.0	53.4	-1.60
8:32:00 AM	54.8	53.0	-1.80
8:33:00 AM	54.6	52.8	-1.80
8:34:00 AM	54.7	52.9	-1.80
8:35:00 AM	54.1	52.4	-1.70
8:36:00 AM	53.8	52.1	-1.70
8:37:00 AM	53.4	52.0	-1.40
8:38:00 AM	53.1	51.8	-1.30
8:39:00 AM	53.8	51.4	-2.40
<b>Promedio</b>	<b>56.46</b>	<b>54.61</b>	
<b>Desviación</b>	<b>1.944</b>	<b>1.733</b>	
<b>Max</b>	<b>59.80</b>	<b>56.90</b>	
<b>Min</b>	<b>53.10</b>	<b>51.80</b>	

Certificado No.:

66949

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-10

\*\*\*\*\*

