



Certificado No.:

67067

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-12

TRACKLINK

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
1:06:00 PM	32.08	31.9	-0.18
1:07:00 PM	32.17	31.9	-0.27
1:08:00 PM	32.18	31.8	-0.38
1:09:00 PM	32.24	31.8	-0.44
1:10:00 PM	32.24	31.8	-0.44
1:11:00 PM	32.34	31.7	-0.64
1:12:00 PM	32.18	31.7	-0.48
1:13:00 PM	32.16	31.7	-0.46
1:14:00 PM	32.31	31.7	-0.61
1:15:00 PM	32.45	31.7	-0.75
1:16:00 PM	32.50	31.7	-0.80
1:17:00 PM	32.65	31.6	-1.05
1:18:00 PM	32.63	31.6	-1.03
1:19:00 PM	32.53	31.5	-1.03
1:20:00 PM	32.19	31.5	-0.69
1:21:00 PM	32.62	31.5	-1.12
1:22:00 PM	33.49	31.5	-1.99
1:23:00 PM	33.32	31.5	-1.82
1:24:00 PM	33.07	31.6	-1.47
1:25:00 PM	32.84	31.6	-1.24
1:26:00 PM	32.66	31.7	-0.96
1:27:00 PM	32.42	31.7	-0.72
1:28:00 PM	32.16	31.7	-0.46
1:29:00 PM	31.99	31.8	-0.19
1:30:00 PM	31.86	31.9	0.04
1:31:00 PM	31.52	31.9	0.38
1:32:00 PM	31.26	32.0	0.74
1:33:00 PM	31.15	32.1	0.95
1:34:00 PM	30.84	32.2	1.36
1:35:00 PM	30.79	32.2	1.41
<b>Promedio</b>	<b>32.23</b>	<b>31.80</b>	
<b>Desviación</b>	<b>0.634</b>	<b>0.196</b>	
<b>Max</b>	<b>33.49</b>	<b>32.20</b>	
<b>Min</b>	<b>30.79</b>	<b>31.50</b>	

Certificado No.:

67067

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-12

TRACKLINK

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
1:06:00 PM	67.50	70.50	3.00
1:07:00 PM	67.60	70.40	2.80
1:08:00 PM	67.20	70.10	2.90
1:09:00 PM	67.10	70.00	2.90
1:10:00 PM	67.00	70.00	3.00
1:11:00 PM	66.50	69.80	3.30
1:12:00 PM	66.80	69.90	3.10
1:13:00 PM	66.50	69.80	3.30
1:14:00 PM	66.40	69.50	3.10
1:15:00 PM	66.20	69.40	3.20
1:16:00 PM	66.00	69.10	3.10
1:17:00 PM	66.00	69.00	3.00
1:18:00 PM	65.70	69.00	3.30
1:19:00 PM	65.50	68.80	3.30
1:20:00 PM	65.20	68.90	3.70
1:21:00 PM	65.00	68.50	3.50
1:22:00 PM	64.80	68.40	3.60
1:23:00 PM	64.90	68.10	3.20
1:24:00 PM	64.50	68.00	3.50
1:25:00 PM	64.40	68.00	3.60
1:26:00 PM	64.10	67.80	3.70
1:27:00 PM	63.80	67.90	4.10
1:28:00 PM	63.90	67.70	3.80
1:29:00 PM	63.50	67.50	4.00
1:30:00 PM	63.40	67.40	4.00
1:31:00 PM	63.20	67.30	4.10
1:32:00 PM	63.10	67.00	3.90
1:33:00 PM	63.00	67.00	4.00
1:34:00 PM	62.80	66.90	4.10
1:35:00 PM	63.10	66.20	3.10
<b>Promedio</b>	<b>65.16</b>	<b>68.60</b>	
<b>Desviación</b>	<b>1.528</b>	<b>1.186</b>	
<b>Max</b>	<b>67.60</b>	<b>70.50</b>	
<b>Min</b>	<b>62.80</b>	<b>66.90</b>	

Certificado No.:

67067

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-12

TRACKLINK

