

Certificado No.:

67069

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-12

TRACKLINK

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
3:07:00 PM	31.40	32.00	0.60
3:08:00 PM	31.49	32.00	0.51
3:09:00 PM	31.57	31.90	0.33
3:10:00 PM	31.61	31.90	0.29
3:11:00 PM	31.73	31.90	0.17
3:12:00 PM	31.86	31.80	-0.06
3:13:00 PM	32.00	31.80	-0.20
3:14:00 PM	32.00	31.80	-0.20
3:15:00 PM	32.10	31.80	-0.30
3:16:00 PM	32.18	31.80	-0.38
3:17:00 PM	32.23	31.80	-0.43
3:18:00 PM	32.30	31.70	-0.60
3:19:00 PM	32.44	31.70	-0.74
3:20:00 PM	32.46	31.60	-0.86
3:21:00 PM	32.56	31.60	-0.96
3:22:00 PM	32.60	31.60	-1.00
3:23:00 PM	32.68	31.60	-1.08
3:24:00 PM	32.76	31.60	-1.16
3:25:00 PM	32.91	31.70	-1.21
3:26:00 PM	33.04	31.70	-1.34
3:27:00 PM	33.29	31.80	-1.49
3:28:00 PM	33.23	31.80	-1.43
3:29:00 PM	32.24	31.80	-0.44
3:30:00 PM	31.72	31.90	0.18
3:31:00 PM	31.24	32.00	0.76
3:32:00 PM	30.78	32.00	1.22
3:33:00 PM	30.47	32.10	1.63
3:34:00 PM	30.11	32.20	2.09
3:35:00 PM	29.85	32.30	2.45
3:36:00 PM	29.51	32.30	2.79
Promedio	31.88	31.85	
Desviación	0.964	0.196	
Max	33.29	32.30	
Min	29.51	31.60	

Certificado No.:

67069

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-12

TRACKLINK

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
3:07:00 PM	70.50	72.50	2.00
3:08:00 PM	70.60	72.40	1.80
3:09:00 PM	70.20	72.10	1.90
3:10:00 PM	70.10	72.00	1.90
3:11:00 PM	70.00	72.00	2.00
3:12:00 PM	69.50	71.80	2.30
3:13:00 PM	69.80	71.90	2.10
3:14:00 PM	69.50	71.80	2.30
3:15:00 PM	69.40	71.50	2.10
3:16:00 PM	69.20	71.40	2.20
3:17:00 PM	69.00	71.10	2.10
3:18:00 PM	69.00	71.00	2.00
3:19:00 PM	68.70	71.00	2.30
3:20:00 PM	68.50	70.80	2.30
3:21:00 PM	68.20	70.90	2.70
3:22:00 PM	68.00	70.50	2.50
3:23:00 PM	67.80	70.40	2.60
3:24:00 PM	67.90	70.10	2.20
3:25:00 PM	67.50	70.00	2.50
3:26:00 PM	67.40	70.00	2.60
3:27:00 PM	67.10	69.80	2.70
3:28:00 PM	66.80	69.90	3.10
3:29:00 PM	66.90	69.70	2.80
3:30:00 PM	66.50	69.50	3.00
3:31:00 PM	66.40	69.40	3.00
3:32:00 PM	66.20	69.30	3.10
3:33:00 PM	66.10	69.00	2.90
3:34:00 PM	66.00	69.00	3.00
3:35:00 PM	65.80	68.90	3.10
3:36:00 PM	66.10	68.20	2.10
Promedio	68.16	70.60	
Desviación	1.528	1.186	
Max	70.60	72.50	
Min	65.80	68.90	

Certificado No.:

67069

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-12

TRACKLINK

