

Certificado No.:

67072

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-12

TRACKLINK

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
3:15:00 PM	21.79	21.40	-0.39
3:16:00 PM	21.55	21.40	-0.15
3:17:00 PM	21.36	21.10	-0.26
3:18:00 PM	21.35	21.00	-0.35
3:19:00 PM	21.21	21.00	-0.21
3:20:00 PM	21.14	20.70	-0.44
3:21:00 PM	20.83	20.50	-0.33
3:22:00 PM	20.77	20.30	-0.47
3:23:00 PM	20.62	20.10	-0.52
3:24:00 PM	20.50	19.70	-0.80
3:25:00 PM	20.37	19.90	-0.47
3:26:00 PM	20.25	19.10	-1.15
3:27:00 PM	20.11	19.20	-0.91
3:28:00 PM	19.96	19.80	-0.16
3:29:00 PM	19.86	18.30	-1.56
3:30:00 PM	19.78	18.70	-1.08
3:31:00 PM	19.62	18.80	-0.82
3:32:00 PM	19.61	18.90	-0.71
3:33:00 PM	19.48	19.10	-0.38
3:34:00 PM	19.33	19.20	-0.13
3:35:00 PM	19.21	19.40	0.19
3:36:00 PM	19.11	19.50	0.39
3:37:00 PM	18.93	19.10	0.17
3:38:00 PM	18.91	18.80	-0.11
3:39:00 PM	18.71	18.20	-0.51
3:40:00 PM	18.65	17.80	-0.85
3:41:00 PM	18.57	17.70	-0.87
3:42:00 PM	18.49	17.80	-0.69
3:43:00 PM	18.35	18.40	0.05
3:44:00 PM	18.26	18.80	0.54
Promedio	19.89	19.46	
Desviación	1.054	1.090	
Max	21.79	21.40	
Min	18.26	17.70	

Certificado No.:

67072

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-12

TRACKLINK

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
3:15:00 PM	56.77	58.90	2.13
3:16:00 PM	56.87	58.80	1.93
3:17:00 PM	56.47	58.50	2.03
3:18:00 PM	56.37	58.40	2.03
3:19:00 PM	56.27	58.40	2.13
3:20:00 PM	55.77	58.20	2.43
3:21:00 PM	56.07	58.30	2.23
3:22:00 PM	55.77	58.20	2.43
3:23:00 PM	55.67	57.90	2.23
3:24:00 PM	55.47	57.80	2.33
3:25:00 PM	55.27	57.50	2.23
3:26:00 PM	55.27	57.40	2.13
3:27:00 PM	54.97	57.40	2.43
3:28:00 PM	54.77	57.20	2.43
3:29:00 PM	54.47	57.30	2.83
3:30:00 PM	54.27	56.90	2.63
3:31:00 PM	54.07	56.80	2.73
3:32:00 PM	54.17	56.50	2.33
3:33:00 PM	53.77	56.40	2.63
3:34:00 PM	53.67	56.40	2.73
3:35:00 PM	53.37	56.20	2.83
3:36:00 PM	53.07	56.30	3.23
3:37:00 PM	53.17	56.10	2.93
3:38:00 PM	52.77	55.90	3.13
3:39:00 PM	52.67	55.80	3.13
3:40:00 PM	52.47	55.70	3.23
3:41:00 PM	52.37	55.40	3.03
3:42:00 PM	52.27	55.40	3.13
3:43:00 PM	52.07	55.30	3.23
3:44:00 PM	52.37	54.60	2.23
Promedio	54.43	57.00	
Desviación	1.528	1.186	
Max	56.87	58.90	
Min	52.07	55.30	

Certificado No.:

67072


Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-12

TRACKLINK

Curva de Calibración - TEMPERATURA



Curva de Calibración - % HUMEDAD RELATIVA

