



Certificado No.:

67149

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-14

Queclink



UBICACION DE EQUIPOS PATRON

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
8:30:00 AM	12.30	12.50	0.20
8:32:00 AM	12.10	12.60	0.50
8:34:00 AM	12.00	12.50	0.50
8:36:00 AM	12.00	12.40	0.40
8:38:00 AM	11.90	12.30	0.40
8:40:00 AM	11.80	12.10	0.30
8:42:00 AM	11.70	12.00	0.30
8:44:00 AM	11.50	11.80	0.30
8:46:00 AM	11.00	11.60	0.60
8:48:00 AM	11.30	11.40	0.10
8:50:00 AM	11.00	11.60	0.60
8:52:00 AM	11.50	11.70	0.20
8:54:00 AM	11.90	11.30	-0.60
8:56:00 AM	12.00	11.20	-0.80
8:58:00 AM	12.40	11.00	-1.40
9:00:00 AM	12.70	11.00	-1.70
9:02:00 AM	12.90	11.40	-1.50
9:04:00 AM	13.00	11.70	-1.30
9:06:00 AM	13.20	11.80	-1.40
9:08:00 AM	13.40	11.90	-1.50
9:10:00 AM	13.60	12.00	-1.60
9:12:00 AM	13.00	12.40	-0.60
9:14:00 AM	12.80	12.50	-0.30
9:16:00 AM	12.70	12.60	-0.10
9:18:00 AM	12.40	12.80	0.40
9:20:00 AM	12.00	12.90	0.90
9:22:00 AM	12.00	12.10	0.10
9:24:00 AM	11.80	11.80	0.00
9:26:00 AM	11.70	11.50	-0.20
9:28:00 AM	11.40	11.30	-0.10
<b>Promedio</b>	<b>12.17</b>	<b>11.92</b>	
<b>Desviación</b>	<b>0.68</b>	<b>0.54</b>	
<b>Max</b>	<b>13.60</b>	<b>12.90</b>	
<b>Min</b>	<b>11.00</b>	<b>11.00</b>	

Certificado No.:

67149

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-14

Queclink

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
8:30:00 AM	75.40	76.00	0.60
8:32:00 AM	75.30	76.30	1.00
8:34:00 AM	75.20	76.40	1.20
8:36:00 AM	75.00	76.50	1.50
8:38:00 AM	75.00	76.30	1.30
8:40:00 AM	74.80	76.00	1.20
8:42:00 AM	74.40	75.90	1.50
8:44:00 AM	74.30	75.80	1.50
8:46:00 AM	74.10	75.70	1.60
8:48:00 AM	74.00	75.60	1.60
8:50:00 AM	74.00	75.80	1.80
8:52:00 AM	73.80	76.00	2.20
8:54:00 AM	73.60	76.00	2.40
8:56:00 AM	73.50	76.20	2.70
8:58:00 AM	73.10	76.40	3.30
9:00:00 AM	73.00	76.50	3.50
9:02:00 AM	73.00	75.80	2.80
9:04:00 AM	72.90	75.40	2.50
9:06:00 AM	72.80	75.20	2.40
9:08:00 AM	72.10	75.00	2.90
9:10:00 AM	72.00	75.00	3.00
9:12:00 AM	72.00	74.90	2.90
9:14:00 AM	71.80	74.80	3.00
9:16:00 AM	71.70	74.70	3.00
9:18:00 AM	71.60	74.00	2.40
9:20:00 AM	72.00	73.70	1.70
9:22:00 AM	72.00	73.60	1.60
9:24:00 AM	72.40	73.40	1.00
9:26:00 AM	72.80	72.70	-0.10
9:28:00 AM	73.00	72.50	-0.50
<b>Promedio</b>	<b>73.35</b>	<b>75.27</b>	
<b>Desviación</b>	<b>1.20</b>	<b>1.14</b>	
<b>Max</b>	<b>75.40</b>	<b>76.50</b>	
<b>Min</b>	<b>71.60</b>	<b>72.70</b>	

Certificado No.:

67149

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-14

Quealink

