



Certificado No.:

67151

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-14

Queclink



UBICACION DE EQUIPOS PATRON

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

| Tiempo (hh:mm:ss) | Temperatura Patrón (°C) | Temperatura Calibrando (°C) | Error (°C) |
|-------------------|-------------------------|-----------------------------|------------|
| 9:15:00 AM        | 12.30                   | 13.00                       | 0.70       |
| 9:17:00 AM        | 12.20                   | 12.80                       | 0.60       |
| 9:19:00 AM        | 12.10                   | 12.40                       | 0.30       |
| 9:21:00 AM        | 12.00                   | 12.10                       | 0.10       |
| 9:23:00 AM        | 11.80                   | 12.00                       | 0.20       |
| 9:25:00 AM        | 11.40                   | 11.80                       | 0.40       |
| 9:27:00 AM        | 11.50                   | 11.40                       | -0.10      |
| 9:29:00 AM        | 11.10                   | 11.10                       | 0.00       |
| 9:31:00 AM        | 11.00                   | 11.10                       | 0.10       |
| 9:33:00 AM        | 11.00                   | 10.80                       | -0.20      |
| 9:35:00 AM        | 10.90                   | 10.50                       | -0.40      |
| 9:37:00 AM        | 10.80                   | 10.10                       | -0.70      |
| 9:39:00 AM        | 11.00                   | 10.80                       | -0.20      |
| 9:41:00 AM        | 11.40                   | 10.90                       | -0.50      |
| 9:43:00 AM        | 11.20                   | 11.00                       | -0.20      |
| 9:45:00 AM        | 11.70                   | 11.20                       | -0.50      |
| 9:47:00 AM        | 11.80                   | 11.40                       | -0.40      |
| 9:49:00 AM        | 11.90                   | 11.50                       | -0.40      |
| 9:51:00 AM        | 12.00                   | 12.00                       | 0.00       |
| 9:53:00 AM        | 12.10                   | 12.10                       | 0.00       |
| 9:55:00 AM        | 12.20                   | 12.50                       | 0.30       |
| 9:57:00 AM        | 12.40                   | 12.40                       | 0.00       |
| 9:59:00 AM        | 12.50                   | 12.80                       | 0.30       |
| 10:01:00 AM       | 13.00                   | 12.90                       | -0.10      |
| 10:03:00 AM       | 13.10                   | 13.00                       | -0.10      |
| 10:05:00 AM       | 13.20                   | 13.10                       | -0.10      |
| 10:07:00 AM       | 13.50                   | 13.00                       | -0.50      |
| 10:09:00 AM       | 13.10                   | 12.80                       | -0.30      |
| 10:11:00 AM       | 12.80                   | 12.50                       | -0.30      |
| 10:13:00 AM       | 12.40                   | 12.10                       | -0.30      |
| <b>Promedio</b>   | <b>11.98</b>            | <b>11.90</b>                |            |
| <b>Desviación</b> | <b>0.76</b>             | <b>0.87</b>                 |            |
| <b>Max</b>        | <b>13.50</b>            | <b>13.10</b>                |            |
| <b>Min</b>        | <b>10.80</b>            | <b>10.10</b>                |            |

Certificado No.:

67151

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-14

Queclink

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

| Tiempo (hh:mm:ss) | Humedad Relativa Patrón (%HR) | Humedad Relativa Calibrando (%HR) | Error (%HR) |
|-------------------|-------------------------------|-----------------------------------|-------------|
| 9:15:00 AM        | 74.80                         | 75.40                             | 0.60        |
| 9:17:00 AM        | 74.40                         | 75.90                             | 1.50        |
| 9:19:00 AM        | 74.10                         | 76.00                             | 1.90        |
| 9:21:00 AM        | 74.00                         | 76.10                             | 2.10        |
| 9:23:00 AM        | 74.00                         | 76.50                             | 2.50        |
| 9:25:00 AM        | 73.80                         | 76.20                             | 2.40        |
| 9:27:00 AM        | 73.40                         | 76.80                             | 3.40        |
| 9:29:00 AM        | 73.10                         | 76.20                             | 3.10        |
| 9:31:00 AM        | 73.00                         | 76.10                             | 3.10        |
| 9:33:00 AM        | 73.00                         | 76.00                             | 3.00        |
| 9:35:00 AM        | 72.80                         | 75.80                             | 3.00        |
| 9:37:00 AM        | 72.40                         | 75.40                             | 3.00        |
| 9:39:00 AM        | 72.10                         | 75.10                             | 3.00        |
| 9:41:00 AM        | 71.80                         | 75.00                             | 3.20        |
| 9:43:00 AM        | 71.90                         | 74.50                             | 2.60        |
| 9:45:00 AM        | 71.50                         | 74.10                             | 2.60        |
| 9:47:00 AM        | 71.10                         | 74.00                             | 2.90        |
| 9:49:00 AM        | 71.00                         | 73.80                             | 2.80        |
| 9:51:00 AM        | 70.50                         | 73.90                             | 3.40        |
| 9:53:00 AM        | 70.10                         | 73.20                             | 3.10        |
| 9:55:00 AM        | 70.50                         | 73.10                             | 2.60        |
| 9:57:00 AM        | 70.40                         | 72.80                             | 2.40        |
| 9:59:00 AM        | 70.20                         | 72.40                             | 2.20        |
| 10:01:00 AM       | 70.80                         | 72.10                             | 1.30        |
| 10:03:00 AM       | 70.90                         | 72.00                             | 1.10        |
| 10:05:00 AM       | 71.10                         | 71.50                             | 0.40        |
| 10:07:00 AM       | 71.40                         | 71.00                             | -0.40       |
| 10:09:00 AM       | 71.20                         | 71.20                             | 0.00        |
| 10:11:00 AM       | 71.50                         | 71.40                             | -0.10       |
| 10:13:00 AM       | 71.80                         | 71.50                             | -0.30       |
| <b>Promedio</b>   | <b>72.09</b>                  | <b>74.17</b>                      |             |
| <b>Desviación</b> | <b>1.39</b>                   | <b>1.88</b>                       |             |
| <b>Max</b>        | <b>74.80</b>                  | <b>76.80</b>                      |             |
| <b>Min</b>        | <b>70.10</b>                  | <b>71.00</b>                      |             |

Certificado No.:

67151

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-14

Quealink

