



Certificado No.:

67154

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-14

Queclink



UBICACION DE EQUIPOS PATRON

Tabla 1.0: Lecturas de la TEMPERATURA en el área de trabajo (°C)

Tiempo (hh:mm:ss)	Temperatura Patrón (°C)	Temperatura Calibrando (°C)	Error (°C)
8:15:00 AM	14.50	14.10	-0.40
8:17:00 AM	14.80	14.00	-0.80
8:19:00 AM	14.90	14.00	-0.90
8:21:00 AM	14.40	13.80	-0.60
8:23:00 AM	14.40	13.40	-1.00
8:25:00 AM	14.20	13.50	-0.70
8:27:00 AM	14.00	13.20	-0.80
8:29:00 AM	14.00	13.00	-1.00
8:31:00 AM	13.80	12.80	-1.00
8:33:00 AM	13.90	12.70	-1.20
8:35:00 AM	13.20	12.40	-0.80
8:37:00 AM	13.10	12.00	-1.10
8:39:00 AM	13.00	12.10	-0.90
8:41:00 AM	12.50	12.00	-0.50
8:43:00 AM	12.40	12.40	0.00
8:45:00 AM	12.10	12.50	0.40
8:47:00 AM	12.10	12.80	0.70
8:49:00 AM	12.00	13.00	1.00
8:51:00 AM	11.80	13.10	1.30
8:53:00 AM	11.40	13.20	1.80
8:55:00 AM	11.10	12.50	1.40
8:57:00 AM	11.00	12.10	1.10
8:59:00 AM	10.80	12.00	1.20
9:01:00 AM	10.40	11.80	1.40
9:03:00 AM	10.50	11.40	0.90
9:05:00 AM	10.20	11.50	1.30
9:07:00 AM	10.00	11.40	1.40
9:09:00 AM	10.80	11.10	0.30
9:11:00 AM	11.50	10.80	-0.70
9:13:00 AM	11.90	10.90	-1.00
<b>Promedio</b>	<b>12.49</b>	<b>12.52</b>	
<b>Desviación</b>	<b>1.53</b>	<b>0.93</b>	
<b>Max</b>	<b>14.90</b>	<b>14.10</b>	
<b>Min</b>	<b>10.00</b>	<b>10.80</b>	

Certificado No.:

67154

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-14

Queclink

Tabla 2.0: Lecturas de la HUMEDAD RELATIVA en el área de trabajo (%HR)

Tiempo (hh:mm:ss)	Humedad Relativa Patrón (%HR)	Humedad Relativa Calibrando (%HR)	Error (%HR)
8:15:00 AM	66.80	62.10	-4.70
8:17:00 AM	66.20	62.50	-3.70
8:19:00 AM	66.30	62.90	-3.40
8:21:00 AM	66.40	62.50	-3.90
8:23:00 AM	66.10	62.40	-3.70
8:25:00 AM	66.00	62.90	-3.10
8:27:00 AM	65.80	63.00	-2.80
8:29:00 AM	65.40	63.00	-2.40
8:31:00 AM	65.10	63.10	-2.00
8:33:00 AM	65.00	63.20	-1.80
8:35:00 AM	65.00	63.40	-1.60
8:37:00 AM	64.80	62.80	-2.00
8:39:00 AM	64.50	62.90	-1.60
8:41:00 AM	64.10	62.10	-2.00
8:43:00 AM	64.20	62.00	-2.20
8:45:00 AM	64.00	61.80	-2.20
8:47:00 AM	64.30	61.40	-2.90
8:49:00 AM	64.00	61.50	-2.50
8:51:00 AM	63.80	61.90	-1.90
8:53:00 AM	63.50	61.10	-2.40
8:55:00 AM	63.40	61.00	-2.40
8:57:00 AM	63.10	61.60	-1.50
8:59:00 AM	63.20	60.80	-2.40
9:01:00 AM	63.00	60.90	-2.10
9:03:00 AM	63.00	60.20	-2.80
9:05:00 AM	62.80	60.30	-2.50
9:07:00 AM	62.80	60.40	-2.40
9:09:00 AM	62.90	60.80	-2.10
9:11:00 AM	62.30	60.90	-1.40
9:13:00 AM	62.40	60.00	-2.40
<b>Promedio</b>	<b>64.34</b>	<b>61.85</b>	
<b>Desviación</b>	<b>1.34</b>	<b>1.02</b>	
<b>Max</b>	<b>66.80</b>	<b>63.40</b>	
<b>Min</b>	<b>62.30</b>	<b>60.20</b>	

Certificado No.:

67154

Registrador de Temperatura y Humedad

Fecha de Calibración:

2026-03-14

Quealink

